

SAMPLE

検査成績表

<TEST REPORT>

HIOKI
HIOKI E.E. CORPORATION

1/1

品名<Model Name> (抵抗計 <RESISTANCE HiTESTER>)
 形名<Model Number> (3541)
 製造番号<Serial No.> (No. 100212975)
 検査年月日<Test Date> (2010-02-10)
 <YYYY-MM-DD>
 検査条件<Test Conditions> (23.1 °C, 47 %rh)

項目 <Item>	レンジ <Range>	入力 <Input>	(標準器校正値) <Std. Cal. Value>	許容範囲 <Tolerance>	表示値 <Indicated Value>
確度<Accuracy>					
Ω	電源周波数<Supply frequency>60Hz, サンプリング<Sampling>SLOW2				
	20mΩ	19.0000mΩ (18.9790 mΩ)	18.9599 mΩ	~	18.9981 mΩ (18.9824 mΩ)
	(OVC ON)				
	200mΩ	190.000mΩ (189.931 mΩ)	189.740 mΩ	~	190.122 mΩ (189.932 mΩ)
	(OVC ON)				
	測定電流<Measurement current> 1A)				
	200mΩ	190.000mΩ (189.917 mΩ)	189.821 mΩ	~	190.013 mΩ (189.931 mΩ)
	(OVC ON)				
	測定電流<Measurement current> 100mA)				
	2Ω	1900.00mΩ (1898.69 mΩ)	1898.41 mΩ	~	1898.97 mΩ (1898.71 mΩ)
	(OVC ON)				
	20Ω	19.0000 Ω (18.9987 Ω)	18.9967 Ω	~	19.0007 Ω (18.9989 Ω)
	(OVC ON)				
	200Ω	190.000 Ω (189.996 Ω)	189.978 Ω	~	190.014 Ω (189.996 Ω)
	2kΩ	1900.00 Ω (1900.01 Ω)	1899.85 Ω	~	1900.17 Ω (1900.03 Ω)
	20kΩ	19.0000kΩ (18.9995 kΩ)	18.9979 kΩ	~	19.0011 kΩ (18.9997 kΩ)
	100kΩ	19.000kΩ (19.000 kΩ)	18.996 kΩ	~	19.004 kΩ (19.000 kΩ)
	100kΩ	100.000kΩ (100.000 kΩ)	99.990 kΩ	~	100.010 kΩ (100.000 kΩ)
	1MΩ	190.00kΩ (189.99 kΩ)	189.95 kΩ	~	190.03 kΩ (189.99 kΩ)
	1MΩ	1000.00kΩ (999.97 kΩ)	999.87 kΩ	~	1000.07 kΩ (999.97 kΩ)
	10MΩ	1.9000MΩ (1.8999 MΩ)	1.8986 MΩ	~	1.9012 MΩ (1.9000 MΩ)
	10MΩ	10.0000MΩ (9.9988 MΩ)	9.9943 MΩ	~	10.0033 MΩ (9.9987 MΩ)
	100MΩ	19.000MΩ (18.999 MΩ)	18.942 MΩ	~	19.056 MΩ (19.005 MΩ)
	100MΩ	100.000MΩ (100.003 MΩ)	99.783 MΩ	~	100.223 MΩ (100.005 MΩ)
LP Ω	電源周波数<Supply frequency>60Hz, サンプリング<Sampling>SLOW2				
	2Ω	1900.00mΩ (1898.69 mΩ)	1898.47 mΩ	~	1898.91 mΩ (1898.72 mΩ)
	(OVC ON)				
	20Ω	19.0000 Ω (18.9987 Ω)	18.9965 Ω	~	19.0009 Ω (18.9989 Ω)
	(OVC ON)				
	200Ω	190.000 Ω (189.996 Ω)	189.974 Ω	~	190.018 Ω (189.998 Ω)
	(OVC ON)				
	2kΩ	1900.00 Ω (1900.01 Ω)	1899.79 Ω	~	1900.23 Ω (1900.04 Ω)
	(OVC ON)				
°C(Pt)	25.0 °C	(25.1 °C)	24.9 °C	~	25.3 °C (25.1 °C)
	75.0 °C	(75.1 °C)	74.9 °C	~	75.3 °C (75.1 °C)
°C(Analog.In)	190.0 °C		187.8 °C	~	192.2 °C (190.0 °C)
(0V/0°C, 1V/100°C)	(1.9V)				
	10.0 °C		9.6 °C	~	10.4 °C (10.0 °C)
	(0.1V)				

備考<Note>

- ・標準器校正値を使用しているポイントの許容範囲は、標準器校正値を基準に定めています。
 <The tolerance for each point using the standard calibration value is based on the standard calibration value.>
- ・FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected by>	承認者<Approved by>
(PASS)	()	()

No.3541-6