



Certificate: 80008921

Master Contract: 300342

Project: 80296539

Date Issued: 2026-06-11

Standards Used	Description
CSA C22.2 No. 61010-2-034:18	Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 2-034: Particular requirements for measurement equipment for insulation resistance and test equipment for electric strength (Adopted IEC 61010-2-34:2017, first edition, with Canadian deviations)
UL 61010-2-034, 1st Edition (2020)	UL Standard for Safety Requirements for Electrical Equipment for Measurement, Control and Laboratory Use – Part 2-034: Particular Requirements for Measurement Equipment for Insulation Resistance and Test Equipment for Electric Strength - First Edition



Certificate: 80008921


Master Contract: 300342

Project: 80296539

Date Issued: 2026-06-11

Notes:

Products certified under Class(es) C363106, C363186 have been certified under CSA's ISO/IEC 17065 accreditation with the Standards Council of Canada (SCC). www.scc.ca





Supplement to Certificate of Compliance

Certificate: 80008921

Master Contract: 300342

*The products listed, including the latest revision described below,
are eligible to be marked in accordance with the referenced Certificate.*

Product Certification History

Project	Date	Description
80296539	2026-06-11	Update report#80008921 to add alternate components.
80159135	2023-03-27	Update report 80008921 to add alternate low-power transformer T102.
80134884	2022-08-11	Update report: 80008921 to add alternate parts.
80008921	2020-01-30	New Model Certification for Insulation Tester ST5520