

SAMPLE

検査成績表
<TEST REPORT>HIOKI
HIOKI E.E. CORPORATION

1/2

品名<Model Name> (マルチファンクションモジュール<MULTIFUNCTION MODULE>)
形名<Model Number> (2306)
製造番号<Serial No.> (No. 090209366)
検査年月日<Test Date> (2009-02-10)
(<YYYY-MM-DD>)
検査条件<Test Conditions> (22.0 °C, 45 %rh)

項目 <Item>	チャンネル <CH>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>	測定値 <Measured Value>
--------------	---------------	----------------	---------------	---------------------	-------------------------

直流電圧
<DC Voltage>

CH1	50 mV	0 mV	-0.040 mV ~	0.040 mV	(0.002 mV)
		40 mV	39.860 mV ~	40.140 mV	(40.000 mV)
	500 mV	0 mV	-0.10 mV ~	0.10 mV	(0.00 mV)
		400 mV	398.90 mV ~	401.10 mV	(400.00 mV)
	5 V	0 V	-0.0010 V ~	0.0010 V	(0.0000 V)
		4 V	3.9890 V ~	4.0110 V	(4.0002 V)
	50 V	0 V	-0.010 V ~	0.010 V	(-0.001 V)
		40 V	39.890 V ~	40.110 V	(39.999 V)

直流電流
<DC Current>

			チャンネル <CH>		
CH1 CH2 CH3 CH4 CH5 CH6 CH7 CH8	30 mA	0 mA	-0.010 mA ~	0.010 mA	1 (0.000 mA)
			-0.010 mA ~	0.010 mA	2 (0.000 mA)
			-0.010 mA ~	0.010 mA	3 (-0.001 mA)
			-0.010 mA ~	0.010 mA	4 (-0.001 mA)
			-0.010 mA ~	0.010 mA	5 (0.000 mA)
			-0.010 mA ~	0.010 mA	6 (0.000 mA)
			-0.010 mA ~	0.010 mA	7 (0.000 mA)
			-0.010 mA ~	0.010 mA	8 (0.000 mA)
CH1 CH2 CH3 CH4 CH5 CH6 CH7 CH8	30 mA	20 mA	19.940 mA ~	20.060 mA	1 (20.001 mA)
			19.940 mA ~	20.060 mA	2 (20.000 mA)
			19.940 mA ~	20.060 mA	3 (20.000 mA)
			19.940 mA ~	20.060 mA	4 (20.000 mA)
			19.940 mA ~	20.060 mA	5 (20.000 mA)
			19.940 mA ~	20.060 mA	6 (19.999 mA)
			19.940 mA ~	20.060 mA	7 (20.000 mA)
			19.940 mA ~	20.060 mA	8 (19.999 mA)

総合判定<Overall Result>	検査者<Inspected by>	承認者<Approved by>
(PASS)	()	()

SAMPLE

検査成績表

<TEST REPORT>

HIOKI
HIOKI E. E. CORPORATION

2/2

製造番号<Serial No.> (No. 090209366)
 検査年月日<Test Date> (2009-02-10)
 <YYYY-MM-DD>

項目 <Item>	チャンネル <CH>	レンジ <Range>	温度 <Temperature>	許容範囲 <Tolerance>	測定値 <Measured Value>
熱電対 *1 <Thermocouple>					
			入力 *2 <Input>		
CH1	K	0 °C	(0mV)	-1.0 °C ~ 1.0 °C	(0.0 °C)
	K	800 °C	(33.275mV)	797.0 °C ~ 803.0 °C	(800.0 °C)
	E	0 °C	(0mV)	-1.0 °C ~ 1.0 °C	(0.0 °C)
	E	500 °C	(37.005mV)	497.8 °C ~ 502.2 °C	(500.0 °C)
	R	400 °C	(3.408mV)	397.0 °C ~ 403.0 °C	(400.0 °C)
熱電対 *3 <Thermocouple>					
			校正值 <Calibration>		
CH4	K	0°C	(-0.3 °C)	-2.3 °C ~ 1.7 °C	(-0.2 °C)
CH6	K	0°C	(-0.2 °C)	-2.2 °C ~ 1.8 °C	(-0.2 °C)
3線式測温抵抗体 <3 wires Resistance Thermometer Sensor>					
			入力 *2 <Input>	チャンネル <CH>	
CH1	Pt100	0 °C	(100 Ω)	-0.5 °C ~ 0.5 °C	1(0.1 °C)
CH2				-0.5 °C ~ 0.5 °C	2(0.1 °C)
CH3				-0.5 °C ~ 0.5 °C	3(0.1 °C)
CH4				-0.5 °C ~ 0.5 °C	4(0.1 °C)
CH5				-0.5 °C ~ 0.5 °C	5(0.1 °C)
CH6				-0.5 °C ~ 0.5 °C	6(0.1 °C)
CH7				-0.5 °C ~ 0.5 °C	7(0.1 °C)
CH8				-0.5 °C ~ 0.5 °C	8(0.1 °C)
CH1	Pt100	500 °C	(280.98 Ω)	498.3 °C ~ 501.7 °C	1(500.1 °C)
CH2				498.3 °C ~ 501.7 °C	2(500.1 °C)
CH3				498.3 °C ~ 501.7 °C	3(500.1 °C)
CH4				498.3 °C ~ 501.7 °C	4(500.1 °C)
CH5				498.3 °C ~ 501.7 °C	5(500.1 °C)
CH6				498.3 °C ~ 501.7 °C	6(500.1 °C)
CH7				498.3 °C ~ 501.7 °C	7(500.1 °C)
CH8				498.3 °C ~ 501.7 °C	8(500.1 °C)

備考<Note>

*1. 基準接点補償:外部
 <Reference junction compensation : External>

*2. 入力温度に相当する疑似入力で行っています。
 <A voltage equivalent to the indicated temperature is used for input during inspection.>

*3. 基準接点補償:内部
 <Reference junction compensation : Internal>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>