

SAMPLE

検査成績表
<TEST REPORT>HIOKI
HIOKI E. E. CORPORATION

1/1

品名<Model Name> (波形モジュール<WAVEFORM MODULE>)
形名<Model Number> (2321)
製造番号<Serial No.> (No. 110498626)
検査年月日<Test Date> (2011-04-25)
(<YYYY-MM-DD>)
検査条件<Test Conditions> (23.3 °C, 53 %rh)

項目 <Item>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>	測定値 <Measured Value>
--------------	----------------	---------------	---------------------	-------------------------

DC精度<DC Accuracy>

				CH1	CH2
50 mV	40 mV	39.75 mV ~	40.25 mV (	40.10 mV) (	40.00 mV)
100 mV	80 mV	79.50 mV ~	80.50 mV (	80.10 mV) (	79.90 mV)
200 mV	160 mV	159.00 mV ~	161.00 mV (	160.20 mV) (	160.10 mV)
500 mV	400 mV	397.5 mV ~	402.5 mV (	400.6 mV) (	400.3 mV)
1 V	0.8 V	0.7950 V ~	0.8050 V (	0.7994 V) (	0.7994 V)
2 V	1.6 V	1.5900 V ~	1.6100 V (	1.6025 V) (	1.6013 V)
5 V	4 V	3.975 V ~	4.025 V (	4.006 V) (	3.997 V)
10 V	8 V	7.950 V ~	8.050 V (	8.006 V) (	7.994 V)
20 V	16 V	15.900 V ~	16.100 V (	16.013 V) (	15.988 V)
50 V	40 V	39.75 V ~	40.25 V (	39.97 V) (	39.97 V)

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected by>	承認者<Approved by>
(PASS)	()	()

No. 2321-2