

SAMPLE

# 検査成績表

&lt;TEST REPORT&gt;

品名<Model Name> ( 電エスタ <MULTI TESTER> )  
形名<Model Number> ( 3008 )  
製造番号<Serial No.> ( No. 110712345 )  
検査年月日<Test Date> ( 2011-07-21 )  
<YYYY-MM-DD>  
検査条件<Test Conditions> ( 23.0 °C, 50 %rh )  
検査周波数<Test Frequency> ( 55Hz )

1/1

| 項目<br><Item> | レンジ<br><Range> | 計器指示値<br><Indicated Value> | 許容範囲<br><Tolerance>   | 標準器指示値<br><Standard> |
|--------------|----------------|----------------------------|-----------------------|----------------------|
| 確度<Accuracy> |                |                            |                       |                      |
| DC V         |                |                            |                       |                      |
| DC           | 6 V            | 6 V                        | 5.850 V ~ 6.150 V (   | 6.000 V )            |
|              | 30 V           | 30 V                       | 29.25 V ~ 30.75 V (   | 30.00 V )            |
|              | 60 V           | 60 V                       | 58.50 V ~ 61.50 V (   | 60.00 V )            |
|              | 300 V          | 300 V                      | 292.5 V ~ 307.5 V (   | 300.0 V )            |
|              | 600 V          | 600 V                      | 585.0 V ~ 615.0 V (   | 600.0 V )            |
| AC V         |                |                            |                       |                      |
| AC           | 6 V            | 6 V                        | 5.850 V ~ 6.150 V (   | 6.000 V )            |
|              | 30 V           | 30 V                       | 29.25 V ~ 30.75 V (   | 30.00 V )            |
|              | 150 V          | 150 V                      | 146.3 V ~ 153.7 V (   | 150.0 V )            |
|              | 300 V          | 300 V                      | 292.5 V ~ 307.5 V (   | 300.0 V )            |
|              | 600 V          | 600 V                      | 585.0 V ~ 615.0 V (   | 600.0 V )            |
| OHMS         |                |                            |                       |                      |
|              | R× 1           | 100 Ω                      | 88.67 Ω ~ 112.7 Ω (   | 100.00 Ω )           |
|              | R× 10          | 1 kΩ                       | 0.886 kΩ ~ 1.127 kΩ ( | 1.000 kΩ )           |
|              | R× 100         | 10 kΩ                      | 8.867 kΩ ~ 11.27 kΩ ( | 10.000 kΩ )          |

備考&lt;Note&gt;

FAIL判定箇所は、グレー表示としています。&lt;FAIL decision points are highlighted in gray.&gt;

| 総合判定<Overall Result> | 検査者<Inspected by> | 承認者<Approved by> |
|----------------------|-------------------|------------------|
| ( PASS )             | ( )               | ( )              |