

SAMPLE

検査成績表

<TEST REPORT>

品名<Model Name> (ハイテスタ <HiTESTER>) 1/1
 形名<Model Number> (3030-01)
 製造番号<Serial No.> (No. 110712345)
 検査年月日<Test Date> (2011-12-31)
 (<YYYY-MM-DD>)
 検査条件<Test Conditions> (23.0 °C, 50 %rh)
 検査周波数<Test Frequency> (55Hz)

項目 <Item>	レンジ <Range>	計器指示値 <Indicated Value>	許容範囲 <Tolerance>	標準器指示値 <Standard>
確度 <Accuracy>				
DC V				
DC	0.3 V	0.3 V	0.292 V ~ 0.307 V	(0.300 V)
	3 V	3 V	2.925 V ~ 3.075 V	(3.000 V)
	12 V	12 V	11.70 V ~ 12.30 V	(12.00 V)
	30 V	30 V	29.25 V ~ 30.75 V	(30.00 V)
	120 V	120 V	117.0 V ~ 123.0 V	(120.0 V)
	300 V	300 V	292.5 V ~ 307.5 V	(300.0 V)
	1000 V	1000 V	975.0 V ~ 1025.0 V	(1000.0 V)
AC V				
AC	12 V	12 V	11.52 V ~ 12.48 V	(12.00 V)
	30 V	30 V	29.25 V ~ 30.75 V	(30.00 V)
	120 V	120 V	117.0 V ~ 123.0 V	(120.0 V)
	300 V	300 V	292.5 V ~ 307.5 V	(300.0 V)
	1000 V	1000 V	975.0 V ~ 1025.0 V	(1000.0 V)
DC A				
	50 μA	50 μA	48.50 μA ~ 51.50 μA	(50.00 μA)
	30 mA	30 mA	29.10 mA ~ 30.90 mA	(30.00 mA)
	300 mA	300 mA	291.0 mA ~ 309.0 mA	(300.0 mA)
OHMS				
	R× 1	30 Ω	26.60 Ω ~ 33.82 Ω	(30.00 Ω)
	R× 10	300 Ω	266.0 Ω ~ 338.2 Ω	(300.0 Ω)
	kΩ	10 kΩ	8.867 kΩ ~ 11.27 kΩ	(10.000 kΩ)
BATT				
DC	1.5V	1.5 V	0.90 V ~ 1.80 V	(1.50 V)

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected By>	承認者<Approved By>
(PASS)	()	()