

SAMPLE 検査成績表  
 <TEST REPORT>

品名<Model Name> ( AC クランプメータ<AC CLAMP METER> )  
 形名<Model Number> ( CM4142 )  
 製造番号<Serial No.> ( No.190112345 )  
 検査年月日<Test Date> ( 2019-01-30 )  
 ( <YYYY-MM-DD> )  
 検査条件<Test Conditions> ( 23.6°C, 54%rh )

項目 <Item>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>	校正値 <Calibration Value>
交流電流<AC Current>	60.00A	5.00A 50Hz	4.85 A ~ 5.15 A	( 5.01 A )
		54.00A 50Hz	53.11 A ~ 54.89 A	( 53.99 A )
	600.0A	20.0A 50Hz	19.4 A ~ 20.6 A	( 20.0 A )
		540.0A 50Hz	531.6 A ~ 548.4 A	( 539.8 A )
	2000A	200A 50Hz	194 A ~ 206 A	( 200 A )
		1800A 50Hz	1770 A ~ 1830 A	( 1800 A )
交流電流Peak<AC Current Peak>	60.00A	120.0Apeak 50Hz	117.4 A ~ 122.6 A	( 120.0 A )
		-120.0Apeak 50Hz	-122.6 A ~ -117.4 A	( -120.0 A )
	600.0A	1200Apeak 50Hz	1179 A ~ 1221 A	( 1200 A )
		-1200Apeak 50Hz	-1221 A ~ -1179 A	( -1199 A )
	2000A	2500Apeak 50Hz	2433 A ~ 2567 A	( 2499 A )
		-2500Apeak 50Hz	-2567 A ~ -2433 A	( -2498 A )
交流電流周波数特性<AC Current Frequency Characteristic>	600.0A Filter ON	200.0A 30Hz	194.5 A ~ 205.5 A	( 199.8 A )
		200.0A 66Hz	195.7 A ~ 204.3 A	( 199.6 A )
	600.0A Filter OFF	200.0A 30Hz	195.5 A ~ 204.5 A	( 199.9 A )
		200.0A 500Hz	195.5 A ~ 204.5 A	( 200.0 A )
		200.0A 1000Hz	195.5 A ~ 204.5 A	( 199.7 A )
電流周波数<AC Current Frequency>	99.99Hz	20.00A 90.00Hz	89.90 Hz ~ 90.10 Hz	( 90.01 Hz )
	999.9Hz	20.00A 900.0Hz	899.0 Hz ~ 901.0 Hz	( 900.0 Hz )
突入電流<Inrush Current>	60.00A Inrush rms	5.00A 50Hz	4.62 A ~ 5.38 A	( 5.04 A )
		54.00A 50Hz	51.17 A ~ 56.83 A	( 53.99 A )
	600.0A Inrush rms	20.0A 50Hz	17.7 A ~ 22.3 A	( 20.1 A )
		540.0A 50Hz	511.7 A ~ 568.3 A	( 539.8 A )
	2000A Inrush rms	200A 50Hz	177 A ~ 223 A	( 199 A )
		1800A 50Hz	1697 A ~ 1903 A	( 1800 A )
突入電流Peak<Inrush Current Peak>*1	60.00A Inrush peak	120.0Apeak 50Hz	113.0 A ~ 127.0 A	( 120.0 A )
	600.0A Inrush peak	1200Apeak 50Hz	1130 A ~ 1270 A	( 1202 A )
	2000A Inrush peak	2500Apeak 50Hz	2275 A ~ 2725 A	( 2500 A )

## 備考&lt;Note&gt;

・FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

\*1 突入電流Peakの校正値は絶対値で表記しています。

<The calibration values for Inrush Current Peak are expressed in terms of absolute value.>

\*2 「直流電圧／交流＋直流電圧」は測定回路共通のため、直流電圧のみ検査しています。

<Because the [DC V / AC+DC V] functions share a common measurement circuit, inspection will only be conducted for DC V. >

\*3 「交流＋直流電圧」は非該当。<AC+DC Voltage range not available. >

総合判定<Overall Result>	検査者<Inspected By>	承認者<Approved By>
( PASS )	( )	( )

SAMPLE

## 検査成績表

&lt;TEST REPORT&gt;

**HIOKI**  
 HIOKI E. E. CORPORATION

品名<Model Name> ( AC クランプメータ<AC CLAMP METER> )  
 形名<Model Number> ( CM4142 )  
 製造番号<Serial No.> ( No.190112345 )  
 検査年月日<Test Date> ( 2019-01-30 )  
 <YYYY-MM-DD>

項目 <Item>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>	校正値 <Calibration Value>
交流電圧<AC Voltage>				
6.000V Filter OFF		0.300V 50Hz	0.295 V ~ 0.305 V	( 0.300 V )
		5.400V 50Hz	5.349 V ~ 5.451 V	( 5.400 V )
60.00V Filter OFF		3.00V 50Hz	2.95 V ~ 3.05 V	( 3.00 V )
		54.00V 50Hz	53.49 V ~ 54.51 V	( 54.00 V )
600.0V Filter OFF		30.0V 50Hz	29.5 V ~ 30.5 V	( 30.0 V )
		540.0V 50Hz	534.9 V ~ 545.1 V	( 539.9 V )
1000V Filter OFF		900V 50Hz	889 V ~ 911 V	( 900 V )
直流電圧／交流＋直流電圧<DC Voltage / AC+DC Voltage> <sup>*2</sup>				
600.0mV <sup>*3</sup>		-540.0mV DC	-543.2 mV ~ -536.8 mV	( -540.0 mV )
		0.0mV DC	-0.5 mV ~ 0.5 mV	( 0.0 mV )
		540.0mV DC	536.8 mV ~ 543.2 mV	( 540.1 mV )
6.000V		-5.400V DC	-5.430 V ~ -5.370 V	( -5.400 V )
		0.000V DC	-0.003 V ~ 0.003 V	( 0.000 V )
		5.400V DC	5.370 V ~ 5.430 V	( 5.400 V )
60.00V		-54.00V DC	-54.30 V ~ -53.70 V	( -54.00 V )
		0.00V DC	-0.03 V ~ 0.03 V	( 0.00 V )
		54.00V DC	53.70 V ~ 54.30 V	( 54.01 V )
600.0V		-540.0V DC	-543.0 V ~ -537.0 V	( -540.0 V )
		0.0V DC	-0.3 V ~ 0.3 V	( 0.0 V )
		540.0V DC	537.0 V ~ 543.0 V	( 540.0 V )
1500V		-900V DC	-907 V ~ -893 V	( -900 V )
		900V DC	893 V ~ 907 V	( 900 V )
交流電圧 Peak<AC Voltage Peak>				
6.000V	±10.00Vpeak	50Hz	9.78 V ~ 10.22 V	( 9.95 V )
			-10.22 V ~ -9.78 V	( -10.06 V )
60.00V	±100.0Vpeak	50Hz	97.8 V ~ 102.2 V	( 100.0 V )
			-102.2 V ~ -97.8 V	( -100.0 V )
600.0V	±1000Vpeak	50Hz	978 V ~ 1022 V	( 999 V )
			-1022 V ~ -978 V	( -1000 V )
直流電圧／交流＋直流電圧 Peak<DC Voltage / AC+DC Voltage Peak> <sup>*2</sup>				
600.0mV <sup>*3</sup>	±1000mVpeak	50Hz	983 mV ~ 1017 mV	( 1001 mV )
			-1017 mV ~ -983 mV	( -1001 mV )
6.000V	±10.00Vpeak	50Hz	9.83 V ~ 10.17 V	( 10.02 V )
			-10.17 V ~ -9.83 V	( -9.99 V )
60.00V	±100.0Vpeak	50Hz	98.3 V ~ 101.7 V	( 100.1 V )
			-101.7 V ~ -98.3 V	( -100.0 V )
600.0V	±1000Vpeak	50Hz	983 V ~ 1017 V	( 1000 V )
			-1017 V ~ -983 V	( -1001 V )
交流電圧周波数特性<AC Voltage Frequency Characteristic>				
6.000V Filter ON		5.400V 15Hz	5.287 V ~ 5.513 V	( 5.387 V )
		5.400V 66Hz	5.322 V ~ 5.478 V	( 5.387 V )
6.000V Filter OFF		5.400V 15Hz	5.314 V ~ 5.486 V	( 5.387 V )
		5.400V 1000Hz	5.314 V ~ 5.486 V	( 5.392 V )
60.00V Filter OFF		54.00V 1000Hz	53.14 V ~ 54.86 V	( 53.83 V )
600.0V Filter OFF		540.0V 1000Hz	531.4 V ~ 548.6 V	( 538.0 V )
交流＋直流電圧 周波数特性<AC+DC Voltage Frequency Characteristic>				
6.000V Filter ON		5.400V 66Hz	5.306 V ~ 5.494 V	( 5.388 V )
6.000V Filter OFF		5.400V 1000Hz	5.306 V ~ 5.494 V	( 5.403 V )
60.00V Filter OFF		54.00V 1000Hz	53.06 V ~ 54.94 V	( 53.90 V )
600.0V Filter OFF		540.0V 1000Hz	531.2 V ~ 548.8 V	( 538.5 V )

SAMPLE

## 検査成績表

&lt;TEST REPORT&gt;

**HIOKI**  
 HIOKI E. E. CORPORATION

品名<Model Name> ( AC クランプメータ<AC CLAMP METER> )  
 形名<Model Number> ( CM4142 )  
 製造番号<Serial No.> ( No.190112345 )  
 検査年月日<Test Date> ( 2019-01-30 )  
 <YYYY-MM-DD>

項目 <Item>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>	校正値 <Calibration Value>
電圧周波数<Voltage Frequency>	9.999Hz	0.600V 9.000Hz	8.988 Hz ~ 9.012 Hz	( 9.000 Hz)
	99.99Hz	0.600V 90.00Hz	89.90 Hz ~ 90.10 Hz	( 90.00 Hz)
	999.9Hz	0.600V 900.0Hz	899.0 Hz ~ 901.0 Hz	( 900.0 Hz)
導通チェック<Continuity Check>	600.0Ω	0.0Ω	-0.5 Ω ~ 0.5 Ω	( 0.0 Ω)
		540.0Ω	535.8 Ω ~ 544.2 Ω	( 540.1 Ω)
抵抗<Resistance>	600.0Ω	0.0Ω	-0.5 Ω ~ 0.5 Ω	( 0.0 Ω)
		540.0Ω	535.8 Ω ~ 544.2 Ω	( 540.2 Ω)
	6.000kΩ	0.000kΩ	-0.005 kΩ ~ 0.005 kΩ	( 0.000 kΩ)
		5.400kΩ	5.358 kΩ ~ 5.442 kΩ	( 5.402 kΩ)
	60.00kΩ	0.00kΩ	-0.05 kΩ ~ 0.05 kΩ	( 0.00 kΩ)
		54.00kΩ	53.58 kΩ ~ 54.42 kΩ	( 54.02 kΩ)
	600.0kΩ	0.0kΩ	-0.5 kΩ ~ 0.5 kΩ	( 0.0 kΩ)
		540.0kΩ	535.8 kΩ ~ 544.2 kΩ	( 540.3 kΩ)
ダイオード<Diode>	1.800V	1.500V	1.485 V ~ 1.515 V	( 1.501 V)
静電容量<Capacitance>	1.000μF	0.009μF	0.004 μF ~ 0.014 μF	( 0.009 μF)
		0.090μF	0.084 μF ~ 0.096 μF	( 0.090 μF)
		0.900μF	0.878 μF ~ 0.922 μF	( 0.900 μF)
	10.00μF	9.00μF	8.78 μF ~ 9.22 μF	( 9.01 μF)
	100.0μF	90.0μF	87.8 μF ~ 92.2 μF	( 90.0 μF)
	1000μF	900μF	878 μF ~ 922 μF	( 900 μF)
温度<Temperature>	-40.0~400.0℃	0.0℃	-3.0 ℃ ~ 3.0 ℃	( -0.3 ℃)
		360.0℃	355.2 ℃ ~ 364.8 ℃	( 359.5 ℃)