

品名<Model Name> (メグオームハイテスタ <MΩ HITESTER>)
形名<Model Number> (3110-04)
製造番号<Serial No.> (No. 161202123)
検査年月日<Test Date> (2016-12-02)
(<YYYY-MM-DD>)
検査条件<Test Conditions> (25.0 °C, 55 %rh)
交流試験周波数<AC Test Frequency> (55Hz)

項目 <Item>	レンジ <Range>	計器指示値 <Indicated Value>	許容範囲 <Tolerance>	標準器指示値 <Standard>
確度<Accuracy>				
MΩ		1000 MΩ	900 MΩ ~	1100 MΩ (1000 MΩ)
		500 MΩ	475.0 MΩ ~	525.0 MΩ (500.0 MΩ)
		200 MΩ	190.0 MΩ ~	210.0 MΩ (200.0 MΩ)
		100 MΩ	95.0 MΩ ~	105.0 MΩ (105.0 MΩ)
		20 MΩ	19.00 MΩ ~	21.00 MΩ (21.00 MΩ)
		10 MΩ	9.50 MΩ ~	10.50 MΩ (10.50 MΩ)
		1 MΩ	0.950 MΩ ~	1.050 MΩ (1.050 MΩ)
ACV	500 V	450.0 V	~	550.0 V (550.0 V)

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected By>	承認者<Approved By>
(PASS)	()	()