

SAMPLE

検査成績表

<TEST REPORT>

1/2

品名<Model Name> (デジタルメガオームハイテスタ<DIGITAL MΩ HiTESTER>)
形名<Model Number> (3116)
製造番号<Serial No.> (No. 110712345)
検査年月日<Test Date> (2011-12-31)
(<YYYY-MM-DD>)
検査条件<Test Conditions> (23.0 °C, 50 %rh)

項目 <Item>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>	表示値 <Indicated Value>
確度<Accuracy> MΩ	250V MODE	0 MΩ	0 MΩ ~ 0.008 MΩ	(0.000 MΩ)
		1.000 MΩ	0.972 MΩ ~ 1.028 MΩ	(1.000 MΩ)
		1.600 MΩ	1.560 MΩ ~ 1.640 MΩ	(1.600 MΩ)
		10.00 MΩ	9.72 MΩ ~ 10.28 MΩ	(10.00 MΩ)
		16.00 MΩ	15.60 MΩ ~ 16.40 MΩ	(16.00 MΩ)
		100.0 MΩ	97.2 MΩ ~ 102.8 MΩ	(100.0 MΩ)
		160.0 MΩ	151.2 MΩ ~ 168.8 MΩ	(168.0 MΩ)
		1000 MΩ	942 MΩ ~ 1058 MΩ	(1000 MΩ)
		1900 MΩ	1797 MΩ ~ 1999 MΩ	(1999 MΩ)
	500V MODE	0 MΩ	0 MΩ ~ 0.008 MΩ	(0.000 MΩ)
		1.000 MΩ	0.972 MΩ ~ 1.028 MΩ	(1.000 MΩ)
		1.600 MΩ	1.560 MΩ ~ 1.640 MΩ	(1.600 MΩ)
		10.00 MΩ	9.72 MΩ ~ 10.28 MΩ	(10.00 MΩ)
		16.00 MΩ	15.60 MΩ ~ 16.40 MΩ	(16.00 MΩ)
		100.0 MΩ	97.2 MΩ ~ 102.8 MΩ	(102.0 MΩ)
		160.0 MΩ	156.0 MΩ ~ 164.0 MΩ	(164.0 MΩ)
		1000 MΩ	942 MΩ ~ 1058 MΩ	(1000 MΩ)
		1900 MΩ	1797 MΩ ~ 1999 MΩ	(1999 MΩ)
	1000V MODE	0 MΩ	0 MΩ ~ 0.008 MΩ	(0.000 MΩ)
		1.000 MΩ	0.972 MΩ ~ 1.028 MΩ	(1.000 MΩ)
		1.600 MΩ	1.560 MΩ ~ 1.640 MΩ	(1.600 MΩ)
		10.00 MΩ	9.72 MΩ ~ 10.28 MΩ	(10.00 MΩ)
		16.00 MΩ	15.60 MΩ ~ 16.40 MΩ	(16.00 MΩ)
		100.0 MΩ	97.2 MΩ ~ 102.8 MΩ	(102.0 MΩ)
		160.0 MΩ	156.0 MΩ ~ 164.0 MΩ	(164.0 MΩ)
		1000 MΩ	972 MΩ ~ 1028 MΩ	(1000 MΩ)
		1900 MΩ	1835 MΩ ~ 1965 MΩ	(1965 MΩ)

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>

(PASS)

検査者<Inspected by>

()

承認者<Approved by>

()

SAMPLE

検査成績表

<TEST REPORT>

2/2

製造番号<Serial No.> (No. 110712345)
 検査年月日<Test Date> (2011-12-31)
 <YYYY-MM-DD>

項目 <Item>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>	表示値 <Indicated Value>
確度<Accuracy>				
AC V		0 V	0 V ~ 6 V	(0 V)
		100 V	93 V ~ 107 V	(100 V)
		600 V	588 V ~ 612 V	(600 V)
OHMS		0 Ω	0 Ω ~ 0.4 Ω	(0.0 Ω)
		100.0 Ω	98.8 Ω ~ 101.2 Ω	(100.0 Ω)
		160.0 Ω	158.4 Ω ~ 161.6 Ω	(160.0 Ω)
		1.000 kΩ	0.988 kΩ ~ 1.012 kΩ	(1.000 kΩ)
		1.900 kΩ	1.881 kΩ ~ 1.919 kΩ	(1.900 kΩ)

項目 <Item>	レンジ <Range>	負荷抵抗 <Load Resistance>	許容範囲 <Tolerance>	標準器指示値 <Standard>
定格測定電流 <Rated Measurement Current>				
1000V MODE		1 MΩ	1.000 mA ~ 1.100 mA	(1.000 mA)