

SAMPLE

検査成績表

<TEST REPORT>

HIOKI
HIOKI E. E. CORPORATION

1/1

品名<Model Name> (フィールドハイテストパック<FIELD HiTEST PACK>)
 形名<Model Number> (3124)
 製造番号<Serial No.> (No. 110712345)
 検査年月日<Test Date> (2011-07-12)
 <YYYY-MM-DD>
 検査条件<Test Conditions> (23.0 °C, 50 %rh)
 交流試験周波数<AC Test Frequency> (55Hz)

項目 <Item>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>	表示値 <Indicated Value>
精度<Accuracy>				
ACV	AC 520V	480 V	472.2 V ~ 487.8 V	(480.0 V)
		260 V	256.1 V ~ 263.9 V	(260.0 V)
		130 V	128.1 V ~ 131.9 V	(130.0 V)
		30 V	29.10 V ~ 30.90 V	(30.00 V)
DCV	DC 300V	300 V	291.0 V ~ 309.0 V	(309.0 V)
		30 V	29.10 V ~ 30.90 V	(30.90 V)
		3 V	2.910 V ~ 3.090 V	(3.000 V)
OHMS	10 Ω	10 Ω	8.52 Ω ~ 11.08 Ω	(8.52 Ω)
ER	1000 Ω	500 Ω	470 Ω ~ 530 Ω	(500 Ω)
		100 Ω	47.0 Ω ~ 53.0 Ω	(53.0 Ω)
		10 Ω	4.70 Ω ~ 5.30 Ω	(4.70 Ω)
M Ω	250V/50M Ω	5 M Ω	4.750 M Ω ~ 5.250 M Ω	(5.000 M Ω)
		0.5 M Ω	0.475 M Ω ~ 0.525 M Ω	(0.500 M Ω)
	500V/100M Ω	10 M Ω	9.50 M Ω ~ 10.50 M Ω	(10.00 M Ω)
		1 M Ω	0.950 M Ω ~ 1.050 M Ω	(1.000 M Ω)

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected by>	承認者<Approved by>
(PASS)	()	()

No.3124-3