

SAMPLE

検査成績表

<TEST REPORT>

HIOKI
HIOKI E.E. CORPORATION

1/1

品名<Model Name> (クランプオンハイテスタ <CLAMP ON HiTESTER>)
形名<Model Number> (3128-10)
製造番号<Serial No.> (No 110712345)
検査年月日<Test Date> (2011-12-31)
(<YYYY-MM-DD>)
検査条件<Test Conditions> (23.0 °C, 50 %rh)
検査周波数<Test Frequency> (55Hz)

項目 <Item>	レンジ <Range>	計器指示値 <Indicated Value>	許容範囲 <Tolerance>	標準器指示値 <Standard>
確 度 <Accuracy>	AC 1500 A	1500 A	1455 A ~ 1545 A	(1500 A)
	600 A	600 A	582.0 A ~ 618.0 A	(600.0 A)
	150 A	150 A	145.5 A ~ 154.5 A	(150.0 A)
	60 A	60 A	58.2 A ~ 61.8 A	(60.0 A)
	15 A	15 A	14.55 A ~ 15.45 A	(15.00 A)
	AC 750 V	750 V	727.5 V ~ 772.5 V	(750.0 V)
	300 V	300 V	291.0 V ~ 309.0 V	(300.0 V)
	150 V	150 V	145.5 V ~ 154.5 V	(150.0 V)
	DC 75 V	75 V	72.75 V ~ 77.25 V	(75.00 V)
	× 100 Ω	3 kΩ	2.66 kΩ ~ 3.38 kΩ	(3.00 kΩ)
	× 1 Ω	30 Ω	26.6 Ω ~ 33.8 Ω	(30.0 Ω)

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected by>	承認者<Approved by>
(PASS)	()	()

No.312810-2