

**SAMPLE**

検査成績表  
<TEST REPORT>

1/1

品名<Model Name> ( 検相器<PHASE DETECTOR> )  
 形名<Model Number> ( 3129 )  
 製造番号<Serial No.> ( No. 110712345 )  
 検査年月日<Test Date> ( 2011-12-31 )  
 ( <YYYY-MM-DD> )  
 検査条件<Test Conditions> ( 23.0 °C, 50 %rh )

項目 <Item>	周波数 <Frequency>	判定基準 <Standard of Judgment>	測定結果 <Measured Value>
検相電圧 < Voltage Detection Level>	AC 45Hz	70V以下にて検相動作 <Detection occurs when test voltage is at or below 70V.>	( 69 V)

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result> ( PASS )	検査者<Inspected by> ( )	承認者<Approved by> ( )
----------------------------------	--------------------------	-------------------------