

検査成績表

<TEST REPORT>

1/1

品名<Model Name> (アーステスタ <EARTH TESTER>)
 形名<Model Number> (3150)
 製造番号<Serial No.> (No. 161202123)
 検査年月日<Test Date> (2016-12-02)
 <YYYY-MM-DD>
 検査条件<Test Conditions> (25.0 °C, 55 %rh)
 検査周波数<Test Frequency> (55 Hz)

項目 ＜Item＞	レンジ ＜Range＞	計器指示値 ＜Indicate Value＞	許容範囲 ＜Tolerance＞				標準器指示値 ＜Standard Value＞				
確度 ＜Accuracy＞	3電極法 ＜3-Pole method＞										
OHMS	× 100	1 Ω	75	Ω	～	125	Ω	(	100	Ω)	
		5 Ω	475	Ω	～	525	Ω	(	500	Ω)	
		10 Ω	975	Ω	～	1025	Ω	(	1000	Ω)	
	× 10	1 Ω	7.5	Ω	～	12.5	Ω	(	10.0	Ω)	
		5 Ω	47.5	Ω	～	52.5	Ω	(	50.0	Ω)	
		10 Ω	97.5	Ω	～	102.5	Ω	(	100.0	Ω)	
	× 1	1 Ω	0.75	Ω	～	1.25	Ω	(	1.00	Ω)	
		5 Ω	4.75	Ω	～	5.25	Ω	(	5.00	Ω)	
		10 Ω	9.75	Ω	～	10.25	Ω	(	10.00	Ω)	
2電極法 ＜2-Pole method＞											
OHMS	× 100	1 Ω	75	Ω	～	125	Ω	(	100	Ω)	
		5 Ω	475	Ω	～	525	Ω	(	500	Ω)	
		10 Ω	975	Ω	～	1025	Ω	(	1000	Ω)	
	× 10	1 Ω	7.5	Ω	～	12.5	Ω	(	10.0	Ω)	
		5 Ω	47.5	Ω	～	52.5	Ω	(	50.0	Ω)	
		10 Ω	97.5	Ω	～	102.5	Ω	(	100.0	Ω)	
AC VOLT		30.0 V	29.1	V	～	30.9	V	(	30.0	V)	

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected By>	承認者<Approved By>
(PASS)	()	()