

SAMPLE

検査成績表

<TEST REPORT>

1/1

品名<Model Name> (アースハイテスタ<EARTH HiTESTER>)
 形名<Model Number> (3151)
 製造番号<Serial No.> (No. 110712345)
 検査年月日<Test Date> (2011-07-22)
 <YYYY-MM-DD>
 検査条件<Test Conditions> (23.0 °C, 50 %rh)
 検査周波数<Test Frequency> (60 Hz)

項目 <Item>	レンジ <Range>	計器指示値 <Indicated Value>	許容範囲 <Tolerance>	標準器指示値 <Standard>
確度 <Accuracy>	3電極法 <3-pole method>			
	× 1	1 Ω	0.75 Ω ~ 1.25 Ω	(1.00 Ω)
		5 Ω	4.75 Ω ~ 5.25 Ω	(5.00 Ω)
		10 Ω	9.75 Ω ~ 10.25 Ω	(10.00 Ω)
	× 10	1 Ω	7.5 Ω ~ 12.5 Ω	(10.0 Ω)
		5 Ω	47.5 Ω ~ 52.5 Ω	(50.0 Ω)
		10 Ω	97.5 Ω ~ 102.5 Ω	(100.0 Ω)
	× 100	1 Ω	75 Ω ~ 125 Ω	(100 Ω)
		5 Ω	475 Ω ~ 525 Ω	(500 Ω)
		10 Ω	975 Ω ~ 1025 Ω	(1000 Ω)
	2電極法 <2-pole method>			
	× 10	1 Ω	7.5 Ω ~ 12.5 Ω	(10.0 Ω)
		5 Ω	47.5 Ω ~ 52.5 Ω	(50.0 Ω)
		10 Ω	97.5 Ω ~ 102.5 Ω	(100.0 Ω)
	× 100	1 Ω	75 Ω ~ 125 Ω	(100 Ω)
		5 Ω	475 Ω ~ 525 Ω	(500 Ω)
		10 Ω	975 Ω ~ 1025 Ω	(1000 Ω)
	AC VOLT	30.0 V	29.1 V ~ 30.9 V	(30.0 V)

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected by>	承認者<Approved by>
(PASS)	()	()