

SAMPLE

## 検査成績表

&lt;TEST REPORT&gt;

1/1

品名<Model Name> ( モータ・ハーモニックハイテスタ<MOTOR/HARMONIC HiTESTER> )  
形名<Model Number> ( 3194 )  
製造番号<Serial No.> ( No. 070317039 )  
検査年月日<Test Date> ( 2007-03-12 )  
( <YYYY-MM-DD> )  
検査条件<Test Conditions> ( 22.1 °C, 39 %rh )

| 項目<br><Item>          | 検査点<br><Test point> | 許容範囲<br><Tolerance>       | チャンネル<br><channel> | 表示値<br><Indicated value> |
|-----------------------|---------------------|---------------------------|--------------------|--------------------------|
| 確度<Accuracy>          |                     |                           |                    |                          |
| D/A出力<br><D/A OUTPUT> | +5V                 | +4. 99000 V ~ +5. 01000 V | 1ch                | ( +4.99651 V)            |
|                       |                     | +4. 99000 V ~ +5. 01000 V | 2ch                | ( +4.99773 V)            |
|                       |                     | +4. 99000 V ~ +5. 01000 V | 3ch                | ( +4.99813 V)            |
|                       |                     | +4. 99000 V ~ +5. 01000 V | 4ch                | ( +4.99849 V)            |
|                       |                     | +4. 99000 V ~ +5. 01000 V | 5ch                | ( +4.99777 V)            |
|                       |                     | +4. 99000 V ~ +5. 01000 V | 6ch                | ( +4.99656 V)            |
|                       |                     | +4. 99000 V ~ +5. 01000 V | 7ch                | ( +4.99864 V)            |
|                       |                     | +4. 99000 V ~ +5. 01000 V | 8ch                | ( +4.99749 V)            |
| 周波数測定<br><Frequency>  | 55Hz                | 54. 944 Hz ~ 55. 056 Hz   | fa                 | ( 55.000 Hz)             |
|                       |                     | 54. 944 Hz ~ 55. 056 Hz   | fb                 | ( 55.000 Hz)             |
|                       |                     | 54. 944 Hz ~ 55. 056 Hz   | fc                 | ( 55.000 Hz)             |
|                       | 100kHz              | 99. 89 kHz ~ 100. 11 kHz  | fa                 | ( 100.00 kHz)            |
|                       |                     | 99. 89 kHz ~ 100. 11 kHz  | fb                 | ( 100.00 kHz)            |
|                       |                     | 99. 89 kHz ~ 100. 11 kHz  | fc                 | ( 100.00 kHz)            |

## 備考&lt;Note&gt;

3194標準装備の9605-01検査成績表は別添付となります。

<Although Model 9605-01 is equipped as a standard function in Model 3194,  
the inspection data sheet for Model 9605-01 is issued separately from that of Model 3194.>

FAIL判定箇所は、グレー表示としています。&lt;FAIL decision points are highlighted in gray.&gt;

| 総合判定<Overall Result> | 検査者<Inspected by> | 承認者<Approved by> |
|----------------------|-------------------|------------------|
| ( PASS )             | ( )               | ( )              |

No. 3194-2