

SAMPLE 検査成績表
<TEST REPORT>

1/1

品名<Model Name> (デジタルクランプハイテスタ<DIGITAL CLAMP HiTESTER>)
 形名<Model Number> (3206)
 製造番号<Serial No.> (No. 110712345)
 検査年月日<Test Date> (2011-12-30)
 <YYYY-MM-DD>
 検査条件<Test Conditions> (23.0 °C, 51 %rh)
 検査周波数<Test Frequency> (50Hz)

項目 <Item>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>	表示値 <Indicated Value>
確度<Accuracy>				
AC V	1000 V	1000 V	982 V ~ 1018 V	1000 V
	1000 V	500 V	489 V ~ 511 V	500 V
	200 V	190 V	187.0 V ~ 193.0 V	190.0 V
	200 V	90 V	88.0 V ~ 92.0 V	90.0 V
AC A	1000 A	900 A	884 A ~ 916 A	900 A
	1000 A	500 A	489 A ~ 511 A	500 A
	200 A	190 A	187.0 A ~ 193.0 A	190.0 A
	200 A	90 A	88.0 A ~ 92.0 A	90.0 A
	20 A	19 A	18.70 A ~ 19.30 A	19.00 A
	20 A	9 A	8.80 A ~ 9.20 A	9.00 A
Ω	2000 Ω	1900 Ω	1874 Ω ~ 1926 Ω	1900 Ω
	2000 Ω	900 Ω	884 Ω ~ 916 Ω	900 Ω
	200 Ω	190 Ω	187.4 Ω ~ 192.6 Ω	190.0 Ω
	200 Ω	90 Ω	88.4 Ω ~ 91.6 Ω	90.0 Ω

備考<Note>
 FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected by>	承認者<Approved by>
(PASS)	()	()

No.3206-2