

SAMPLE

検査成績表

<TEST REPORT>

1/1

品名<Model Name> (ペンシルハイテスタ<PENCIL HiTESTER>)
 形名<Model Number> (3218)
 製造番号<Serial No.> (No. 110712345)
 検査年月日<Test Date> (2011-12-31)
 <YYYY-MM-DD>
 検査条件<Test Conditions> (23.0 °C, 50 %rh)
 検査周波数<Test Frequency> (50Hz)

項目 <Item>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>	表示値 <Indicated Value>
確度<Accuracy>				
DC V	200 mV	190.0 mV	185.8 mV ~ 194.2 mV	190.0 mV
	2 V	1.900 V	1.883 V ~ 1.917 V	1.900 V
	20 V	19.00 V	18.72 V ~ 19.28 V	19.00 V
	200 V	190.0 V	187.2 V ~ 192.8 V	190.0 V
	500 V	500 V	490 V ~ 510 V	500 V
AC V	2 V	1.900 V	1.849 V ~ 1.951 V	1.900 V
	20 V	19.00 V	18.49 V ~ 19.51 V	19.00 V
	200 V	190.0 V	184.9 V ~ 195.1 V	190.0 V
	500 V	500 V	481 V ~ 519 V	500 V
Ω	200 Ω	190.0 Ω	185.8 Ω ~ 194.2 Ω	190.0 Ω
	2 KΩ	1.900 KΩ	1.858 KΩ ~ 1.942 KΩ	1.900 KΩ
	20 KΩ	19.00 KΩ	18.58 KΩ ~ 19.42 KΩ	19.00 KΩ
	200 KΩ	190.0 KΩ	185.8 KΩ ~ 194.2 KΩ	190.0 KΩ
	2000 KΩ	1900 KΩ	1858 KΩ ~ 1942 KΩ	1900 KΩ
	20 MΩ	18.00 MΩ	16.16 MΩ ~ 19.84 MΩ	18.00 MΩ

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected by>	承認者<Approved by>
(PASS)	()	()
No.3218-4		