

SAMPLE

検査成績表

<TEST REPORT>

HIOKI
HIOKI E.E. CORPORATION

1/1

品名<Model Name> (リペアハイテスタ<REPAIR HiTESTER>)
形名<Model Number> (3228)
製造番号<Serial No.> (No. 110712345)
検査年月日<Test Date> (2011-12-31)
(<YYYY-MM-DD>)
検査条件<Test Conditions> (25.0 °C, 50 %rh)
検査周波数 (55 Hz)
<Test Frequency>

項目 <Item>	レンジ <Range>		入力 <Input>		許容範囲 <Tolerance>				表示値 <Indication>		
確度<Accuracy>											
DC V	10	V	9.000	V	8.970	V	~	9.030	V (	9.000	V)
	25	V	24.00	V	23.90	V	~	24.10	V (	24.00	V)
DC A	20	mA	18.00	mA	17.80	mA	~	18.20	mA (	18.00	mA)
	200	mA	180.0	mA	178.0	mA	~	182.0	mA (	180.0	mA)
	1	A	0.900	A	0.879	A	~	0.921	A (	0.900	A)
F	100	%	50	%	44	%	~	56	% (	50	%)
OHMS	100	mΩ	90.00	mΩ	89.45	mΩ	~	90.55	mΩ (	90.00	mΩ)
	1	Ω	900.0	Ω	896.9	Ω	~	903.1	Ω (	900.0	Ω)
	10	Ω	9.000	Ω	8.969	Ω	~	9.031	Ω (	9.000	Ω)
	100	Ω	90.00	Ω	89.51	Ω	~	90.49	Ω (	90.00	Ω)
	1	kΩ	900.0	Ω	895.1	Ω	~	904.9	Ω (	900.0	Ω)
	10	kΩ	9.000	kΩ	8.951	kΩ	~	9.049	kΩ (	9.000	kΩ)
	100	kΩ	90.00	kΩ	89.51	kΩ	~	90.49	kΩ (	90.00	kΩ)
	1	MΩ	450.0	kΩ	447.4	kΩ	~	452.6	kΩ (	450.0	kΩ)
	1	MΩ	900.0	kΩ	890.6	kΩ	~	909.4	kΩ (	900.0	kΩ)

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected by>	承認者<Approved by>
(PASS)	()	()

No.3228-3