

SAMPLE

検査成績表

<TEST REPORT>

1/1

品名<Model Name> (プリントハイテスタ<PRINT HiTESTER>)
 形名<Model Number> (3234)
 製造番号<Serial No.> (No. 110712345)
 検査年月日<Test Date> (2011-12-31)
 (<YYYY-MM-DD>)
 検査条件<Test Conditions> (23.0 °C, 50 %rh)
 検査周波数<Test Frequency> (50Hz)

項目 <Item>	ファンクション <Function>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>		表示値 <Indicated Value>
確度 <Accuracy>	DCV	300 mV	300.0 mV	298.8 mV	~ 301.2 mV	(300.0 mV)
		3 V	3.000 V	2.988 V	~ 3.012 V	(3.000 V)
		30 V	30.00 V	29.83 V	~ 30.17 V	(30.00 V)
		300 V	300.0 V	298.3 V	~ 301.7 V	(300.0 V)
		500 V	500 V	495 V	~ 505 V	(500 V)
	ACV	3 V	3.000 V	2.966 V	~ 3.034 V	(3.000 V)
		30 V	30.00 V	29.66 V	~ 30.34 V	(30.00 V)
		300 V	300.0 V	296.6 V	~ 303.4 V	(300.0 V)
		500 V	500 V	491 V	~ 509 V	(500 V)
	DCA	300 mA	300.0 mA	296.8 mA	~ 303.2 mA	(300.0 mA)
	ACA	300 mA	300.0 mA	296.0 mA	~ 304.0 mA	(300.0 mA)
	Ω	300 Ω	280.0 Ω	278.4 Ω	~ 281.6 Ω	(280.0 Ω)
		3 kΩ	2.800 kΩ	2.784 kΩ	~ 2.816 kΩ	(2.800 kΩ)
		30 kΩ	28.00 kΩ	27.84 kΩ	~ 28.16 kΩ	(28.00 kΩ)
		300 kΩ	280.0 kΩ	278.4 kΩ	~ 281.6 kΩ	(280.0 kΩ)
		3000 kΩ	2800 kΩ	2770 kΩ	~ 2830 kΩ	(2800 kΩ)
		30 MΩ	28.00 MΩ	27.42 MΩ	~ 28.58 MΩ	(28.00 MΩ)
	LP Ω	3 kΩ	2.800 kΩ	2.782 kΩ	~ 2.818 kΩ	(2.800 kΩ)
		30 kΩ	28.00 kΩ	27.82 kΩ	~ 28.18 kΩ	(28.00 kΩ)
		300 kΩ	280.0 kΩ	278.2 kΩ	~ 281.8 kΩ	(280.0 kΩ)
		3000 kΩ	2800 kΩ	2768 kΩ	~ 2832 kΩ	(2800 kΩ)
		30 MΩ	28.00 MΩ	27.40 MΩ	~ 28.60 MΩ	(28.00 MΩ)
	OPTION	DC 300 mV	300.0 mV	2988	~ 3012	(3000)

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected by>	承認者<Approved by>
(PASS)	()	()

No.3234-5