

SAMPLE

検査成績表

<TEST REPORT>

HIOKI
HIOKI E.E. CORPORATION

1/1

品名<Model Name> (デジタルハイテスタ<DIGITAL HiTESTER>)
形名<Model Number> (3242)
製造番号<Serial No.> (No. 110712345)
検査年月日<Test Date> (2011-12-31)
(<YYYY-MM-DD>)
検査条件<Test Conditions> (23.0 °C, 50 %rh)
検査周波数<Test Frequency> (50Hz)

項目 <Item>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>	表示値 <Indicated Value>
確度<Accuracy>				
DCV	200 mV	190.0 mV	187.9 mV ~ 192.1 mV	(190.0 mV)
	2 V	1.900 V	1.883 V ~ 1.917 V	(1.900 V)
	20 V	19.00 V	18.72 V ~ 19.28 V	(19.00 V)
	200 V	190.0 V	187.2 V ~ 192.8 V	(190.0 V)
	500 V	500 V	490 V ~ 510 V	(500 V)
ACV	2 V	1.900 V	1.849 V ~ 1.951 V	(1.900 V)
	20 V	19.00 V	18.49 V ~ 19.51 V	(19.00 V)
	200 V	190.0 V	184.9 V ~ 195.1 V	(190.0 V)
	500 V	500 V	481 V ~ 519 V	(500 V)
OHMS	200 Ω	190.0 Ω	185.8 Ω ~ 194.2 Ω	(190.0 Ω)
	2 kΩ	1.900 kΩ	1.858 kΩ ~ 1.942 kΩ	(1.900 kΩ)
	20 kΩ	19.00 kΩ	18.58 kΩ ~ 19.42 kΩ	(19.00 kΩ)
	200 kΩ	190.0 kΩ	185.8 kΩ ~ 194.2 kΩ	(190.0 kΩ)
	2000 kΩ	1900 kΩ	1858 kΩ ~ 1942 kΩ	(1900 kΩ)
	20 MΩ	18.00 MΩ	16.16 MΩ ~ 19.84 MΩ	(18.00 MΩ)

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>

(PASS)

検査者<Inspected by>

()

承認者<Approved by>

()

No.3242-3