

SAMPLE

検査成績表

<TEST REPORT>

1/2

品名<Model Name> (デジタルマルチメータ)
<DIGITAL MULTIMETER>
形名<Model Number> (DT4224)
製造番号<Serial No.> (No. 170904575)
検査年月日<Test Date> (2017-09-01)
<YYYY-MM-DD>
検査条件<Test Conditions> (22.1 °C, 61 %rh)

項目 <Item>	ファンクション <Function>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>	校正値 *1 <Calibration Value>
確度 <Accuracy>	交流電圧 <AC voltage>	6 V	0.06 V 50Hz	0.052 V ~ 0.068 V	(0.060 V)
			5.4 V 50Hz	5.343 V ~ 5.457 V	(5.399 V)
			5.4 V 50Hz	5.262 V ~ 5.538 V	(5.394 V *2)
			5.4 V 300Hz	5.316 V ~ 5.484 V	(5.400 V *3)
		60 V	0.6 V 50Hz	0.52 V ~ 0.68 V	(0.60 V)
			54 V 50Hz	53.43 V ~ 54.57 V	(53.99 V)
		600 V	6 V 50Hz	5.2 V ~ 6.8 V	(6.0 V)
			540 V 50Hz	534.3 V ~ 545.7 V	(539.9 V)
	直流電圧 <DC voltage>	600 mV	0 mV	-0.5 mV ~ 0.5 mV	(0.0 mV)
			-540 mV	-543.2 mV ~ -536.8 mV	(-539.7 mV)
			540 mV	536.8 mV ~ 543.2 mV	(540.0 mV)
		6 V	0 V	-0.005 V ~ 0.005 V	(0.000 V)
			-5.4 V	-5.432 V ~ -5.368 V	(-5.399 V)
			5.4 V	5.368 V ~ 5.432 V	(5.400 V)
		60 V	0 V	-0.05 V ~ 0.05 V	(0.00 V)
			-54 V	-54.32 V ~ -53.68 V	(-53.99 V)
			54 V	53.68 V ~ 54.32 V	(54.00 V)
		600 V	0 V	-0.5 V ~ 0.5 V	(0.0 V)
			-540 V	-543.2 V ~ -536.8 V	(-540.0 V)
			540 V	536.8 V ~ 543.2 V	(540.0 V)

備考<Note>

*1. FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

*2. フィルタON(100Hz) <Filter ON(100Hz)>

*3. フィルタON(500Hz) <Filter ON(500Hz)>

総合判定<Overall Result>	検査者<Inspected By>	承認者<Approved By>
(PASS)	()	()

検査成績表
＜TEST REPORT＞

2/2

製造番号<Serial No.> (No. 170904575)
検査年月日<Test Date> (2017-09-01)
<YYYY-MM-DD>

項目 <Item>	ファンクション <Function>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>	校正値 *1 <Calibration Value>
確度 <Accuracy>	周波数 <Frequency>	100 Hz 1000 Hz 10 kHz	90 Hz (0.6V) 900 Hz (6V) 9 kHz (60V)	89.89 Hz ~ 90.11 Hz 898.9 Hz ~ 901.1 Hz 8.989 kHz ~ 9.011 kHz	(90.01 Hz) (900.0 Hz) (9.001 kHz)
導通 <Continuity>		600 Ω	0 Ω 540 Ω	-0.5 Ω ~ 0.5 Ω 534.1 Ω ~ 545.9 Ω	(0.0 Ω) (540.7 Ω)
抵抗 <Resistance>		600 Ω	0 Ω 540 Ω	-0.5 Ω ~ 0.5 Ω 534.7 Ω ~ 545.3 Ω	(0.0 Ω) (540.6 Ω)
		6 kΩ	0 kΩ 5.4 kΩ	-0.005 kΩ ~ 0.005 kΩ 5.347 kΩ ~ 5.453 kΩ	(0.000 kΩ) (5.402 kΩ)
		60 kΩ	0 kΩ 54 kΩ	-0.05 kΩ ~ 0.05 kΩ 53.47 kΩ ~ 54.53 kΩ	(0.00 kΩ) (54.02 kΩ)
		600 kΩ	0 kΩ 540 kΩ	-0.5 kΩ ~ 0.5 kΩ 534.7 kΩ ~ 545.3 kΩ	(0.0 kΩ) (540.2 kΩ)
		6 MΩ	0 MΩ 3 MΩ	-0.005 MΩ ~ 0.005 MΩ 2.968 MΩ ~ 3.032 MΩ	(0.000 MΩ) (3.001 MΩ)
		60 MΩ	0 MΩ 30 MΩ	-0.05 MΩ ~ 0.05 MΩ 29.50 MΩ ~ 30.50 MΩ	(0.00 MΩ) (29.97 MΩ)
静電容量 <Electrostatic capacity>		1 μF	0.009 μF 0.09 μF 0.9 μF	0.004 μF ~ 0.014 μF 0.084 μF ~ 0.096 μF 0.878 μF ~ 0.922 μF	(0.006 μF) (0.090 μF) (0.900 μF)
		10 μF	9 μF	8.78 μF ~ 9.22 μF	(9.00 μF)
		100 μF	90 μF	87.8 μF ~ 92.2 μF	(90.0 μF)
		1 mF	0.9 mF	0.878 mF ~ 0.922 mF	(0.900 mF)
		10 mF	9 mF	8.35 mF ~ 9.65 mF	(8.97 mF)
ダイオード <Diode>		1.5 V	0.400 V DC	0.392 V ~ 0.408 V	(0.400 V)