

SAMPLE

検査成績表

HIOKI
HIOKI E. E. CORPORATION

<TEST REPORT>

1/1

品名<Model Name> (ペンシルハイテスタ<PENCIL HiTESTER>)
形名<Model Number> (3246 *)
製造番号<Serial No.> (No. 101023395)
検査年月日<Test Date> (2010-11-16)
(<YYYY-MM-DD>)
検査条件<Test Conditions> (22.0 °C, 49 %rh)

項目 <Item>	ファンクション <Function>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>	表示値 <Indicated value>
確度 <Accuracy>	DCV	420.0 mV	0.0 mV	-0.4 mV ~ 0.4 mV	(0.0 mV)
		420.0 mV	370.0 mV	364.8 mV ~ 375.2 mV	(369.7 mV)
		4.200 V	3.700 V	3.648 V ~ 3.752 V	(3.694 V)
		42.00 V	37.00 V	36.48 V ~ 37.52 V	(36.98 V)
		420.0 V	370.0 V	364.8 V ~ 375.2 V	(370.1 V)
		600 V	520 V	510 V ~ 530 V	(519 V)
	ACV (50Hz)	4.200 V	3.700 V	3.607 V ~ 3.793 V	(3.682 V)
	(50Hz)	42.00 V	37.00 V	36.07 V ~ 37.93 V	(36.89 V)
	(50Hz)	420.0 V	370.0 V	360.7 V ~ 379.3 V	(369.4 V)
	(500Hz)	600 V	520 V	501 V ~ 539 V	(514 V)
	Ω	420.0 Ω	0.0 Ω	0.0 Ω ~ 0.4 Ω	(0.0 Ω)
		420.0 Ω	370.0 Ω	362.2 Ω ~ 377.8 Ω	(367.4 Ω)
		4.200 kΩ	3.700 kΩ	3.622 kΩ ~ 3.778 kΩ	(3.677 kΩ)
		42.00 kΩ	37.00 kΩ	36.22 kΩ ~ 37.78 kΩ	(36.77 kΩ)
		420.0 kΩ	370.0 kΩ	362.2 kΩ ~ 377.8 kΩ	(367.8 kΩ)
		4.200 MΩ	3.700 MΩ	3.511 MΩ ~ 3.889 MΩ	(3.680 MΩ)
		42.00 MΩ	37.00 MΩ	33.26 MΩ ~ 40.74 MΩ	(36.78 MΩ)

備考<Note>

- * 3246と3246-01,3246-02,3246-10,3246-20,3246-60,3246-61,3246-62,3246-65,3246-70は、検査成績表のポイントが共通であるため、代表して形名を3246と表記しています。
<Because the inspection points of Models 3246,3246-01,3246-02,3246-10,3246-20,3246-60,3246-61,3246-62,3246-65 and 3246-70 are the same,"Model 3246" is used in this data sheet to represent all ten models.>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected by>	承認者<Approved by>
(PASS)	()	()

No.3246-5