

SAMPLE

検査成績表

HIOKI
HIOKI E. E. CORPORATION

<TEST REPORT>

1/2

品名<Model Name> (デジタルハイテスタ<DIGITAL HiTESTER>)
形名<Model Number> (3252)
製造番号<Serial No.> (No. 110712345)
検査年月日<Test Date> (2011-12-31)
<YYYY-MM-DD>
検査条件<Test Conditions> (23.0 °C, 50 %rh)
検査周波数<Test Frequency> (50Hz)

項目 <Item>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>	表示値 <Indicated Value>
確度<Accuracy>				
DC V	500 mV	500.0 mV	499.1 mV ~ 500.9 mV	(499.1 mV)
	5 V	5.000 V	4.991 V ~ 5.009 V	(4.991 V)
	50 V	50.00 V	49.91 V ~ 50.09 V	(49.91 V)
	500 V	500.0 V	499.1 V ~ 500.9 V	(499.1 V)
	1000 V	1000 V	995 V ~ 1005 V	(995 V)
DC mV	50 mV	50.00 mV	49.65 mV ~ 50.35 mV	(49.65 mV)
	500 mV	500.0 mV	499.1 mV ~ 500.9 mV	(499.1 mV)
AC V	500 mV	500.0 mV	491.5 mV ~ 508.5 mV	(491.5 mV)
	5 V	5.000 V	4.973 V ~ 5.027 V	(4.973 V)
	50 V	50.00 V	49.73 V ~ 50.27 V	(49.73 V)
	500 V	500.0 V	497.3 V ~ 502.7 V	(497.3 V)
	750 V	750 V	737 V ~ 763 V	(737 V)

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected by>	承認者<Approved by>
(PASS)	()	()

No.3252-5

SAMPLE

検査成績表

HIOKI
HIOKI E. E. CORPORATION

<TEST REPORT>

2/2

製造番号<Serial No.> (No. 110712345)
検査年月日<Test Date> (2011-12-31)
<YYYY-MM-DD>

項目 <Item>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>	表示値 <Indicated Value>
精度<Accuracy>				
DC A	500 μ A	500.0 μ A	497.1 μ A ~ 502.9 μ A	(497.1 μ A)
	5000 μ A	5000 μ A	4973 μ A ~ 5027 μ A	(4973 μ A)
	50 mA	50.00 mA	49.71 mA ~ 50.29 mA	(49.71 mA)
	500 mA	500.0 mA	497.3 mA ~ 502.7 mA	(497.3 mA)
	5 A	5.000 A	4.946 A ~ 5.054 A	(4.946 A)
	10 A	9.00 A	8.85 A ~ 9.15 A	(8.85 A)
AC A	500 μ A	500.0 μ A	489.6 μ A ~ 510.4 μ A	(489.6 μ A)
	5000 μ A	5000 μ A	4923 μ A ~ 5077 μ A	(4923 μ A)
	50 mA	50.00 mA	49.21 mA ~ 50.79 mA	(49.21 mA)
	500 mA	500.0 mA	492.3 mA ~ 507.7 mA	(492.3 mA)
	5 A	5.000 A	4.921 A ~ 5.079 A	(4.921 A)
	10 A	9.00 A	8.80 A ~ 9.20 A	(8.80 A)
FREQUENCY	150 Hz	130.00 Hz	129.97 Hz ~ 130.03 Hz	(129.97 Hz)
	1500 Hz	1300.0 Hz	1299.7 Hz ~ 1300.3 Hz	(1299.7 Hz)
	15 kHz	13.000 kHz	12.997 kHz ~ 13.003 kHz	(12.997 kHz)
	150 kHz	130.00 kHz	129.97 kHz ~ 130.03 kHz	(129.97 kHz)
	400 kHz	400.0 kHz	399.9 kHz ~ 400.1 kHz	(399.9 kHz)
Ω	500 Ω	500.0 Ω	497.1 Ω ~ 502.9 Ω	(497.1 Ω)
	5 k Ω	5.000 k Ω	4.988 k Ω ~ 5.012 k Ω	(4.988 k Ω)
	50 k Ω	50.00 k Ω	49.88 k Ω ~ 50.12 k Ω	(49.88 k Ω)
	500 k Ω	500.0 k Ω	498.8 k Ω ~ 501.2 k Ω	(498.8 k Ω)
	5 M Ω	5.000 M Ω	4.988 M Ω ~ 5.012 M Ω	(4.988 M Ω)
	50 M Ω	50.00 M Ω	49.48 M Ω ~ 50.52 M Ω	(49.48 M Ω)

No.3252-5