

SAMPLE

検査成績表

<TEST REPORT>

品名<Model Name> (アナログユニット<ANALOG UNIT>)
形名<Model Number> (MR8905)
製造番号<Serial No.> (No.200812345)
検査年月日<Test Date> (2020-08-03)
(<YYYY-MM-DD>)
検査条件<Test Conditions> (23.0 °C 50 %rh)

| 項目 <Item> | レンジ <Range> | 入力 <Input> | 許容範囲 <Tolerance> | 校正値<Calibration Value> | |
|---------------------------|----------------|---------------|------------------------|------------------------|----------------|
| | | | | CH1 | CH2 |
| 確度<Accuracy> | | | | | |
| DCモード<DC Mode> | | | | | |
| 500 mV/DIV | 0 V | DC | -0.0500 V ~ 0.0500 V | (0.0000 V) | (0.0000 V) |
| 500 mV/DIV | -10 V | DC | -10.0500 V ~ -9.9500 V | (-10.0000 V) | (-10.0000 V) |
| 500 mV/DIV | 10 V | DC | 9.9500 V ~ 10.0500 V | (10.0000 V) | (10.0000 V) |
| 1 V/DIV | 20 V | DC | 19.9000 V ~ 20.1000 V | (20.0000 V) | (20.0000 V) |
| 2 V/DIV | 40 V | DC | 39.8000 V ~ 40.2000 V | (40.0000 V) | (40.0000 V) |
| 5 V/DIV | 100 V | DC | 99.500 V ~ 100.500 V | (100.000 V) | (100.000 V) |
| 10 V/DIV | 200 V | DC | 199.000 V ~ 201.000 V | (200.000 V) | (200.000 V) |
| 20 V/DIV | 400 V | DC | 398.000 V ~ 402.000 V | (400.000 V) | (400.000 V) |
| 50 V/DIV | 1000 V | DC | 995.00 V ~ 1005.00 V | (1000.00 V) | (1000.00 V) |
| RMSモード<RMS Mode> | | | | | |
| 500 mV/DIV | 10 V | 100Hz | 9.8500 V ~ 10.1500 V | (10.0000 V) | (10.0000 V) |
| 1 V/DIV | 20 V | 100Hz | 19.7000 V ~ 20.3000 V | (20.0000 V) | (20.0000 V) |
| 2 V/DIV | 40 V | 100Hz | 39.4000 V ~ 40.6000 V | (40.0000 V) | (40.0000 V) |
| 5 V/DIV | 100 V | 100Hz | 98.500 V ~ 101.500 V | (100.000 V) | (100.000 V) |
| 10 V/DIV | 200 V | 100Hz | 197.000 V ~ 203.000 V | (200.000 V) | (200.000 V) |
| 20 V/DIV | 400 V | 100Hz | 394.000 V ~ 406.000 V | (400.000 V) | (400.000 V) |
| 50 V/DIV | 600 V | 100Hz | 585.00 V ~ 615.00 V | (600.00 V) | (600.00 V) |

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

| 総合判定<Overall Result> | 検査者<Inspected by> | 承認者<Approved by> |
|----------------------|-------------------|------------------|
| (PASS) | () | () |