

品名<Model Name> (極超絶縁計)
 形名<Model Number> (<ULTRA MEGOHMMETER>)
 製造番号<Production No.> (No. SM-10E)
 検査年月日<Test Date> (123456789)
 (2014-09-09)
 (<YYYY-MM-DD>)
 検査条件<Test Condition> (23.0 °C , 60 %rh)

機能 <Function>	レンジ <Range>	許容範囲 <Tolerance>	*校正値 <Calibration Value>
確度<Accuracy>			
測定電圧 <Measuring Voltage>	10 V	9.70 V ~ 10.30 V	(10.00 V)
	25 V	24.25 V ~ 25.75 V	(25.00 V)
	50 V	48.5 V ~ 51.5 V	(50.0 V)
	100 V	97.0 V ~ 103.0 V	(100.0 V)
	250 V	242.5 V ~ 257.5 V	(250.0 V)
	500 V	485 V ~ 515 V	(500 V)
	1000 V	970 V ~ 1030 V	(1000 V)

機能 <Function>	レンジ <Range>	指示値 <Indicated value>	許容範囲 <Tolerance>	*校正値 <Calibration Value>
測定精度 <Measuring Accuracy>				
×	10 MΩ	10 MΩ	9.0 MΩ ~ 11.0 MΩ	(10.0 MΩ)
×	10 ² MΩ	100 MΩ	90 MΩ ~ 110 MΩ	(100 MΩ)
×	10 ³ MΩ	1000 MΩ	900 MΩ ~ 1100 MΩ	(1000 MΩ)
×	10 ⁴ MΩ	10 GΩ	9.0 GΩ ~ 11.0 GΩ	(10.0 GΩ)
×	10 ⁵ MΩ	100 GΩ	90 GΩ ~ 110 GΩ	(100 GΩ)
×	10 ⁶ MΩ	1000 GΩ	900 GΩ ~ 1100 GΩ	(1000 GΩ)
×	10 ⁷ MΩ	10 TΩ	9.0 TΩ ~ 11.0 TΩ	(10.0 TΩ)
×	10 ⁸ MΩ	100 TΩ	90 TΩ ~ 110 TΩ	(100 TΩ)
×	10 MΩ	100 MΩ	90 MΩ ~ 110 MΩ	(100 MΩ)
×	10 ² MΩ	1000 MΩ	900 MΩ ~ 1100 MΩ	(1000 MΩ)
×	10 ³ MΩ	10 GΩ	9.0 GΩ ~ 11.0 GΩ	(10.0 GΩ)
×	10 ⁴ MΩ	100 GΩ	90 GΩ ~ 110 GΩ	(100 GΩ)
×	10 ⁵ MΩ	1000 GΩ	900 GΩ ~ 1100 GΩ	(1000 GΩ)
×	10 ⁶ MΩ	10 TΩ	9.0 TΩ ~ 11.0 TΩ	(10.0 TΩ)
×	10 ⁷ MΩ	100 TΩ	90 TΩ ~ 110 TΩ	(100 TΩ)
×	10 ⁸ MΩ	1000 TΩ	900 TΩ ~ 1100 TΩ	(1000 TΩ)
×	10 MΩ	500 MΩ	400 MΩ ~ 600 MΩ	(500 MΩ)
×	10 ² MΩ	5000 MΩ	4000 MΩ ~ 6000 MΩ	(5000 MΩ)
×	10 ³ MΩ	50 GΩ	40.0 GΩ ~ 60.0 GΩ	(50.0 GΩ)
×	10 ⁴ MΩ	500 GΩ	400 GΩ ~ 600 GΩ	(500 GΩ)
×	10 ⁵ MΩ	5000 GΩ	4000 GΩ ~ 6000 GΩ	(5000 GΩ)

備考

測定精度は、電圧の擬似入力で実施しています。

<Measuring Accuracy is implementing a voltage by pseudoinput>

* FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray>

総合判定<Judgment>	担当者<Inspected by>	承認者<Approved by>
(PASS)	()	()

品名<Model Name> (極超絶縁計)
 形名<Model Number> (<ULTRA MEGOHMMETER>)
 製造番号<Production No.> (No. SM-10E)
 123456789)

機能 <Function>	レンジ <Range>	指示値 <Indicated value>	公称値 <Nominal value>	許容範囲 <Tolerance>	*校正値 <Calibration Value>
測定精度 <Measuring Accuracy>					
×	10 ⁶ MΩ		50 TΩ	40.0 TΩ ~ 60.0 TΩ (	50.0 TΩ)
×	10 ⁷ MΩ		500 TΩ	400 TΩ ~ 600 TΩ (	500 TΩ)
×	10 ⁸ MΩ		5000 TΩ	4000 TΩ ~ 6000 TΩ (	5000 TΩ)

機能 ＜Function＞	レンジ ＜Range＞	指示値 ＜Indicated value＞	許容範囲 ＜Tolerance＞	*校正値 ＜Calibration Value＞	
抵抗測定 ＜Resistant measurement＞					
1000V	×	10 MΩ	10 MΩ	9.0 MΩ ～ 11.0 MΩ (10.0 MΩ)	
	×	10 MΩ	100 MΩ	90 MΩ ～ 110 MΩ (100 MΩ)	
	×	10 ² MΩ	100 MΩ	90 MΩ ～ 110 MΩ (100 MΩ)	
	×	10 ² MΩ	1000 MΩ	900 MΩ ～ 1100 MΩ (1000 MΩ)	
	×	10 ³ MΩ	1000 MΩ	900 MΩ ～ 1100 MΩ (1000 MΩ)	
	×	10 ³ MΩ	10 GΩ	9.0 GΩ ～ 11.0 GΩ (10.0 GΩ)	
	×	10 ⁴ MΩ	10 GΩ	9.0 GΩ ～ 11.0 GΩ (10.0 GΩ)	
	×	10 ⁴ MΩ	100 GΩ	90 GΩ ～ 110 GΩ (100 GΩ)	
	10V	×	10 ⁵ MΩ	1000 MΩ	900 MΩ ～ 1100 MΩ (1000 MΩ)
		×	10 ⁵ MΩ	10 GΩ	9.0 GΩ ～ 11.0 GΩ (10.0 GΩ)
		×	10 ⁶ MΩ	10 GΩ	9.0 GΩ ～ 11.0 GΩ (10.0 GΩ)
		×	10 ⁶ MΩ	100 GΩ	90 GΩ ～ 110 GΩ (100 GΩ)
		×	10 ⁷ MΩ	100 GΩ	90 GΩ ～ 110 GΩ (100 GΩ)
		×	10 ⁷ MΩ	1000 GΩ	900 GΩ ～ 1100 GΩ (1000 GΩ)
×		10 ⁸ MΩ	1000 GΩ	900 GΩ ～ 1100 GΩ (1000 GΩ)	

絶縁試験 <Insulation Resistance>	*校正値 <Calibration Value>
外筐一電源入力 【DC 1000V、100MΩ以上】 <Case - Power Input> <DC 1000V , 100 M Ω or more>	(200.00 MΩ)
赤色測定棒 【DC 1000V、1TΩ以上】 <Red Measuring Rod> <DC 1000V , 1 T Ω or more>	(1.00 TΩ)

機能 <Function>	結果 <Result>
機構試験 【外観・操作部・端子類を点検し良好なこと】 <Inspect the external appearance, operation part, terminals are without defect>	(PASS)

No. SM10E-2