

SAMPLE

検査成績表

<TEST REPORT>

1/1

品名<Model Name> (デジタルハイテスタ<DIGITAL HiTESTER>)
形名<Model Number> (3255)
製造番号<Serial No.> (No. 110712345)
検査年月日<Test Date> (2011-12-31)
(<YYYY-MM-DD>)
検査条件<Test Conditions> (23.0 °C, 50 %rh)
検査周波数<Test Frequency> (50Hz)

項目 <Item>	ファンクション <Function>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>	表示値 <Indicated Value>
確度 <Accuracy>	DCmV	320.0 mV	300.0 mV	298.1 mV ~ 301.9 mV	(298.1 mV)
	DCV	3.200 V	3.000 V	2.981 V ~ 3.019 V	(2.981 V)
		32.00 V	30.00 V	29.57 V ~ 30.43 V	(29.57 V)
		320.0 V	300.0 V	295.7 V ~ 304.3 V	(300.0 V)
		600 V	500 V	490 V ~ 510 V	(490 V)
	ACmV	320.0 mV	300.0 mV	293.2 mV ~ 306.8 mV	(293.2 mV)
	ACV	3.200 V	3.000 V	2.932 V ~ 3.068 V	(2.932 V)
		32.00 V	30.00 V	29.32 V ~ 30.68 V	(29.32 V)
		320.0 V	300.0 V	293.2 V ~ 306.8 V	(293.2 V)
		600 V	500 V	482 V ~ 518 V	(482 V)
	Ω	320.0 Ω	300.0 Ω	294.9 Ω ~ 305.1 Ω	(294.9 Ω)
		3.200 kΩ	3.000 kΩ	2.951 kΩ ~ 3.049 kΩ	(2.951 kΩ)
		32.00 kΩ	30.00 kΩ	29.51 kΩ ~ 30.49 kΩ	(30.00 kΩ)
		320.0 kΩ	300.0 kΩ	295.1 kΩ ~ 304.9 kΩ	(295.1 kΩ)
		3.200 MΩ	3.000 MΩ	2.951 MΩ ~ 3.049 MΩ	(2.951 MΩ)
		32.00 MΩ	30.00 MΩ	26.96 MΩ ~ 33.04 MΩ	(26.96 MΩ)

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected by>	承認者<Approved by>
(PASS)	()	()

No.3255-4