

SAMPLE

検査成績表

<TEST REPORT>

HIOKI
HIOKI E. E. CORPORATION

1/1

品名<Model Name> (デジタルハイテスタ<DIGITAL HiTESTER>)
形名<Model Number> (3255-50)
製造番号<Serial No.> (No. 070321528)
検査年月日<Test Date> (2007-03-14)
(<YYYY-MM-DD>)
検査条件<Test Conditions> (21.9 °C, 34 %rh)

項目 <Item>	ファンクション <Function>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>	表示値 <Indicated value>
確度					
<Accuracy>	ACV (50Hz)	4.200 V	3.700 V	3.622 V ~ 3.778 V	(3.696 V)
		(50Hz) 42.00 V	37.00 V	36.52 V ~ 37.48 V	(36.99 V)
		(50Hz) 420.0 V	370.0 V	364.1 V ~ 375.9 V	(370.6 V)
		(500Hz) 1000 V	900 V	883 V ~ 917 V	(907 V)
	DCV	4.200 V	0.000 V	-0.004 V ~ 0.004 V	(0.000 V)
		4.200 V	3.700 V	3.659 V ~ 3.741 V	(3.698 V)
		42.00 V	37.00 V	36.78 V ~ 37.22 V	(37.00 V)
		420.0 V	370.0 V	367.1 V ~ 372.9 V	(370.9 V)
		1000 V	900 V	890 V ~ 910 V	(905 V)
	ACmV (50Hz)	420.0 mV	370.0 mV	362.2 mV ~ 377.8 mV	(369.8 mV)
	DCmV	420.0 mV	0.0 mV	-0.4 mV ~ 0.4 mV	(0.0 mV)
		420.0 mV	370.0 mV	365.9 mV ~ 374.1 mV	(370.0 mV)
	CLAMP (50Hz)	1000 A	900 A	878 A ~ 922 A	(899 A)
(上記の電流入力は電圧による疑似入力で検査しています)					
<For the above range, calibration is conducted using a test voltage that simulate the condition corresponding to the specified current input.>					
Ω		420.0 Ω	0.0 Ω	0.0 Ω ~ 0.8 Ω	(0.4 Ω)
		420.0 Ω	370.0 Ω	365.5 Ω ~ 374.5 Ω	(370.0 Ω)
		4.200 kΩ	3.700 kΩ	3.671 kΩ ~ 3.729 kΩ	(3.696 kΩ)
		42.00 kΩ	37.00 kΩ	36.59 kΩ ~ 37.41 kΩ	(36.99 kΩ)
		420.0 kΩ	370.0 kΩ	365.9 kΩ ~ 374.1 kΩ	(370.2 kΩ)
		4.200 MΩ	3.700 MΩ	3.622 MΩ ~ 3.778 MΩ	(3.696 MΩ)
		42.00 MΩ	37.00 MΩ	35.11 MΩ ~ 38.89 MΩ	(37.31 MΩ)

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected by>	承認者<Approved by>
(PASS)	()	()

No.325550-2