

SAMPLE

検査成績表

<TEST REPORT>

1/2

品名<Model Name> (デジタルハイテスタ<DIGITAL HiTESTER>)
形名<Model Number> (3256)
製造番号<Serial No.> (No. 1997-1234567)
検査年月日<Test Date> (2011-10-12)
<YYYY-MM-DD>
検査条件<Test Conditions> (25.3 °C, 45 %rh)
検査周波数<Test Frequency> (50Hz)

項目 <Item>	ファンクション <Function>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>	表示値 <Indicated Value>
確度 <Accuracy>					
DCV		560.0 mV	500.0 mV	497.9 mV ~ 502.1 mV	(499.9 mV)
		5.600 V	5.000 V	4.981 V ~ 5.019 V	(4.999 V)
		56.00 V	50.00 V	49.68 V ~ 50.32 V	(49.92 V)
		560.0 V	500.0 V	496.8 V ~ 503.2 V	(498.9 V)
		1000 V	1000 V	988 V ~ 1012 V	(1000 V)
ACV		560.0 mV	500.0 mV	489.0 mV ~ 511.0 mV	(503.5 mV)
		5.600 V	5.000 V	4.921 V ~ 5.079 V	(5.006 V)
		56.00 V	50.00 V	49.21 V ~ 50.79 V	(49.97 V)
		560.0 V	500.0 V	492.1 V ~ 507.9 V	(499.4 V)
		750 V	700 V	686 V ~ 714 V	(700 V)
Hz		149.99 Hz	140.00 Hz	139.97 Hz ~ 140.03 Hz	(140.00 Hz)
		1499.9 Hz	1400.0 Hz	1399.7 Hz ~ 1400.3 Hz	(1400.0 Hz)
		14.999 kHz	14.000 kHz	13.997 kHz ~ 14.003 kHz	(14.000 kHz)
		149.99 kHz	140.00 kHz	139.97 kHz ~ 140.03 kHz	(140.00 kHz)
		400.0 kHz	400.0 kHz	399.9 kHz ~ 400.1 kHz	(399.9 kHz)
Ω		560.0 Ω	500.0 Ω	494.4 Ω ~ 505.6 Ω	(499.2 Ω)
		5.600 kΩ	5.000 kΩ	4.971 kΩ ~ 5.029 kΩ	(4.991 kΩ)
		56.00 kΩ	50.00 kΩ	49.71 kΩ ~ 50.29 kΩ	(50.02 kΩ)
		560.0 kΩ	500.0 kΩ	497.1 kΩ ~ 502.9 kΩ	(499.5 kΩ)
		5.600 MΩ	5.000 MΩ	4.946 MΩ ~ 5.054 MΩ	(4.986 MΩ)
		56.00 MΩ	50.00 MΩ	48.96 MΩ ~ 51.04 MΩ	(49.78 MΩ)

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected by>	承認者<Approved by>
(PASS)	()	()

No.3256-9

SAMPLE

検査成績表

<TEST REPORT>

2/2

製造番号<Serial No.> (No. 1997-1234567)
検査年月日<Test Date> (2011-10-12)
<YYYY-MM-DD>

項目 <Item>	ファンクション <Function>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>	表示値 <Indicated Value>
精度 <Accuracy>	DCA	56.00 μ A	50.00 μ A	49.21 μ A ~ 50.79 μ A	(50.02 μ A)
		560.0 μ A	500.0 μ A	492.1 μ A ~ 507.9 μ A	(500.3 μ A)
		5600 μ A	5000 μ A	4921 μ A ~ 5079 μ A	(5002 μ A)
		56.00 mA	50.00 mA	49.21 mA ~ 50.79 mA	(50.04 mA)
		560.0 mA	500.0 mA	492.1 mA ~ 507.9 mA	(500.6 mA)
		10.00 A	9.00 A	8.83 A ~ 9.17 A	(8.97 A)
	ACA	56.00 μ A	50.00 μ A	48.71 μ A ~ 51.29 μ A	(50.13 μ A)
		560.0 μ A	500.0 μ A	486.5 μ A ~ 513.5 μ A	(504.4 μ A)
		5600 μ A	5000 μ A	4871 μ A ~ 5129 μ A	(5013 μ A)
		56.00 mA	50.00 mA	48.65 mA ~ 51.35 mA	(50.45 mA)
		560.0 mA	500.0 mA	487.1 mA ~ 512.9 mA	(501.9 mA)
		10.00 A	9.00 A	8.74 A ~ 9.26 A	(8.98 A)
導通チェック <Continuity Check>		560.0 Ω	500.0 Ω	491.9 Ω ~ 508.1 Ω	(497.7 Ω)
ダイオードチェック <Diode Check>		2.000 V	DC 1.900 V	1.803 V ~ 1.997 V	(1.888 V)