

SAMPLE

検査成績表

<TEST REPORT>

HIOKI
HIOKI E. E. CORPORATION

1/2

品名<Model Name> (デジタルハイテスタ<DIGITAL HiTESTER>)
形名<Model Number> (3256-50*1)
製造番号<Serial No.> (No. 091234567)
検査年月日<Test Date> (2011-10-12)
(<YYYY-MM-DD>)
検査条件<Test Conditions> (26.0 °C, 44 %rh)
検査周波数<AC Test Frequency> (50 Hz)

項目 <Item>	ファンクション <Function>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>	表示値 <Indicated value>
確度 <Accuracy>	DCV	420.0 mV	0.0 mV	-0.2 mV ~ 0.2 mV	(0.0 mV)
		420.0 mV	400.0 mV	397.8 mV ~ 402.2 mV	(399.8 mV)
		4.200 V	4.000 V	3.978 V ~ 4.022 V	(4.001 V)
		42.00 V	40.00 V	39.78 V ~ 40.22 V	(40.03 V)
		420.0 V	400.0 V	397.8 V ~ 402.2 V	(400.8 V)
		1000 V	1000 V	993 V ~ 1007 V	(1001 V)
	ACV (50Hz)	420.0 mV	400.0 mV	393.7 mV ~ 406.3 mV	(399.8 mV)
		4.200 V	4.000 V	3.949 V ~ 4.051 V	(3.994 V)
		42.00 V	40.00 V	39.49 V ~ 40.51 V	(40.00 V)
		420.0 V	400.0 V	394.9 V ~ 405.1 V	(400.7 V)
		1000 V	1000 V	982 V ~ 1018 V	(998 V)
	Hz	199.99 Hz	150.00 Hz	149.95 Hz ~ 150.05 Hz	(150.00 Hz)
		1999.9 Hz	1500.0 Hz	1499.6 Hz ~ 1500.4 Hz	(1499.9 Hz)
		19.999 kHz	15.000 kHz	14.996 kHz ~ 15.004 kHz	(14.999 kHz)
		199.99 kHz	150.00 kHz	149.96 kHz ~ 150.04 kHz	(149.99 kHz)
		500.0 kHz	500.0 kHz	499.8 kHz ~ 500.2 kHz	(500.0 kHz)
	Ω	420.0 Ω	0.0 Ω	0.0 Ω ~ 0.4 Ω	(0.2 Ω)
		420.0 Ω	400.0 Ω	396.8 Ω ~ 403.2 Ω	(399.4 Ω)
		4.200 kΩ	4.000 kΩ	3.970 kΩ ~ 4.030 kΩ	(3.997 kΩ)
		42.00 kΩ	40.00 kΩ	39.70 kΩ ~ 40.30 kΩ	(39.98 kΩ)
		420.0 kΩ	400.0 kΩ	397.0 kΩ ~ 403.0 kΩ	(399.8 kΩ)
		4.200 MΩ	4.000 MΩ	3.938 MΩ ~ 4.062 MΩ	(4.003 MΩ)
		42.00 MΩ	40.00 MΩ	38.98 MΩ ~ 41.02 MΩ	(40.01 MΩ)
	導通チェック <Continuity Check>	420.0 Ω	400.0 Ω	396.8 Ω ~ 403.2 Ω	(399.5 Ω)
	ダイオードチェック <Diode Check>	2.00 V	2.00 V	1.88 V ~ 2.12 V	(2.00 V)

総合判定<Overall Result>

(PASS)

検査者<Inspected by>

()

承認者<Approved by>

()

No.325650-7

SAMPLE

検査成績表

HIOKI
HIOKI E. E. CORPORATION

<TEST REPORT>

2/2

製造番号<Serial No.> (No. 091234567)
検査年月日<Test Date> (2011-10-12)
<YYYY-MM-DD>

項目 <Item>	ファンクション <Function>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>	表示値 <Indicated value>
精度 <Accuracy>	DCA	42.00 μ A	40.00 μ A	39.36 μ A ~ 40.64 μ A	(40.00 μ A)
		420.0 μ A	400.0 μ A	393.6 μ A ~ 406.4 μ A	(399.4 μ A)
		4200 μ A	4000 μ A	3936 μ A ~ 4064 μ A	(3999 μ A)
		42.00 mA	40.00 mA	39.36 mA ~ 40.64 mA	(39.95 mA)
		420.0 mA	400.0 mA	393.6 mA ~ 406.4 mA	(399.8 mA)
		10.00 A	9.00 A	8.83 A ~ 9.17 A	(9.00 A)
	ACA	42.00 μ A	40.00 μ A	38.95 μ A ~ 41.05 μ A	(40.00 μ A)
	(50Hz)	420.0 μ A	400.0 μ A	389.5 μ A ~ 410.5 μ A	(399.8 μ A)
		4200 μ A	4000 μ A	3895 μ A ~ 4105 μ A	(4003 μ A)
		42.00 mA	40.00 mA	38.95 mA ~ 41.05 mA	(39.99 mA)
		420.0 mA	400.0 mA	389.5 mA ~ 410.5 mA	(400.3 mA)
		10.00 A	9.00 A	8.73 A ~ 9.27 A	(9.00 A)

備考<Note>

- * 1 3256-50と3256-51は、検査成績表のポイントが共通であるため、代表して形名を3256-50と表記しています。
<Because the inspection points of Models 3256-50 and 3256-51 are the same, "Model 3256-50" is used in this data sheet to represent both models.>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

No.325650-7