

SAMPLE

検査成績表

<TEST REPORT>

品名<Model Name> (デジタルハイテスタ<DIGITAL HiTESTER>)
 形名<Model Number> (3257-50*1)
 製造番号<Serial No.> (No. 090503136)
 検査年月日<Test Date> (2011-05-23)
 (<YYYY-MM-DD>)
 検査条件<Test Conditions> (25.1 °C, 46 %rh)
 検査周波数<AC Test Frequency> (50 Hz)

項目 <Item>	ファンクション <Function>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>	表示値 <Indicated value>
確度					
<Accuracy>	DCV	420.0 mV	0.0 mV	-0.2 mV ~ 0.2 mV	(0.0 mV)
		420.0 mV	400.0 mV	397.8 mV ~ 402.2 mV	(399.9 mV)
		4.200 V	4.000 V	3.978 V ~ 4.022 V	(3.999 V)
		42.00 V	40.00 V	39.78 V ~ 40.22 V	(39.99 V)
		420.0 V	400.0 V	397.8 V ~ 402.2 V	(399.9 V)
		1000 V	1000 V	993 V ~ 1007 V	(999 V)
	ACV	420.0 mV	400.0mV (50Hz)	393.7 mV ~ 406.3 mV	(399.9 mV)
		4.200 V	4.000V (50Hz)	3.949 V ~ 4.051 V	(3.980 V)
		42.00 V	40.00V (50Hz)	39.49 V ~ 40.51 V	(39.81 V)
		42.00 V	40.00V (500Hz)	39.49 V ~ 40.51 V	(40.22 V)
		420.0 V	400.0V (50Hz)	394.9 V ~ 405.1 V	(398.1 V)
		1000 V	1000V (50Hz)	982 V ~ 1018 V	(995 V)
	Hz	199.99 Hz	150.00 Hz	149.95 Hz ~ 150.05 Hz	(150.00 Hz)
		1999.9 Hz	1500.0 Hz	1499.6 Hz ~ 1500.4 Hz	(1500.0 Hz)
		19.999 kHz	15.000 kHz	14.996 kHz ~ 15.004 kHz	(15.000 kHz)
		199.99 kHz	150.00 kHz	149.96 kHz ~ 150.04 kHz	(150.00 kHz)
		500.0 kHz	500.0 kHz	499.8 kHz ~ 500.2 kHz	(500.0 kHz)
	Ω	420.0 Ω	0.0 Ω	0.0 Ω ~ 0.4 Ω	(0.0 Ω)
		420.0 Ω	400.0 Ω	396.8 Ω ~ 403.2 Ω	(399.3 Ω)
		4.200 kΩ	4.000 kΩ	3.970 kΩ ~ 4.030 kΩ	(3.999 kΩ)
		42.00 kΩ	40.00 kΩ	39.70 kΩ ~ 40.30 kΩ	(39.99 kΩ)
		420.0 kΩ	400.0 kΩ	397.0 kΩ ~ 403.0 kΩ	(400.1 kΩ)
		4.200 MΩ	4.000 MΩ	3.938 MΩ ~ 4.062 MΩ	(4.009 MΩ)
		42.00 MΩ	40.00 MΩ	38.98 MΩ ~ 41.02 MΩ	(40.29 MΩ)
	導通チェック <Continuity Check>	420.0 Ω	400.0 Ω	396.8 Ω ~ 403.2 Ω	(400.0 Ω)
	ダイオードチェック <Diode Check>	2.00 V	2.00 V	1.88 V ~ 2.12 V	(2.00 V)

総合判定<Overall Result>

(PASS)

検査者<Inspected by>

()

承認者<Approved by>

()

SAMPLE

検査成績表

HIOKI
HIOKI E. E. CORPORATION

<TEST REPORT>

2/2

製造番号<Serial No.>
検査年月日<Test Date>(No. 090503136)
(2011-05-23)
<YYYY-MM-DD>

項目 <Item>	ファンクション <Function>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>	表示値 <Indicated value>
確度 <Accuracy>	DCA	42.00 uA 420.0 uA 4200 uA 42.00 mA 420.0 mA 10.00 A	40.00 uA 400.0 uA 4000 uA 40.00 mA 400.0 mA 9.00 A	39.36 uA ~ 40.64 uA 393.6 uA ~ 406.4 uA 3936 uA ~ 4064 uA 39.36 mA ~ 40.64 mA 393.6 mA ~ 406.4 mA 8.83 A ~ 9.17 A	(40.02 uA) (400.3 uA) (3999 uA) (39.99 mA) (399.6 mA) (8.99 A)
	ACA	42.00 uA 420.0 uA 4200 uA 42.00 mA 420.0 mA 10.00 A	40.00uA (50Hz) 400.0uA (50Hz) 4000uA (50Hz) 40.00mA (50Hz) 400.0mA (50Hz) 9.00A (50Hz)	38.95 uA ~ 41.05 uA 389.5 uA ~ 410.5 uA 3895 uA ~ 4105 uA 38.95 mA ~ 41.05 mA 389.5 mA ~ 410.5 mA 8.73 A ~ 9.27 A	(40.10 uA) (399.4 uA) (3998 uA) (39.93 mA) (399.5 mA) (9.00 A)
	DUTY	100.0 % 100.0 % 100.0 % 100.0 %	10.0% (10Hz) 10.0% (100Hz) 90.0% (1kHz) 90.0% (10kHz)	8.4 % ~ 11.6 % 8.4 % ~ 11.6 % 87.6 % ~ 92.4 % 84.1 % ~ 95.9 %	(9.7 %) (9.7 %) (90.0 %) (87.3 %)

備考<Note>

- * 1 3257-50と3257-51は、検査成績表のポイントが共通であるため、代表して形名を3257-50と表記しています。
<Because the inspection points of Models 3257-50 and 3257-51 are the same,"Model 3257-50" is used in this data sheet to represent both models.>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

No.325750-6