

SAMPLE

検査成績表

<TEST REPORT>

品名<Model Name> (クランプオンハイテスタ<CLAMP ON HiTESTER>)
形名<Model Number> (3291)
製造番号<Serial No.> (No. 080415059)
検査年月日<Test Date> (2011 - 05- 13)
<YYYY-MM-DD>
検査条件<Test Conditions> (23.7 °C, 43 %rh)

項目 <Item>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>	表示値 <Indicated value>
確度<Accuracy>				
FILTER ON	1000 A	900 A	55Hz 882 A ~ 918 A	(908 A)
	600 A	540 A	55Hz 531.4 A ~ 548.6 A	(543.0 A)
	60 A	54 A	55Hz 53.14 A ~ 54.86 A	(54.24 A)
FILTER OFF	1000 A	900 A	100Hz 868 A ~ 932 A	(904 A)
	600 A	540 A	100Hz 523.3 A ~ 556.7 A	(541.8 A)
	60 A	54 A	100Hz 52.33 A ~ 55.67 A	(54.26 A)

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected by>	承認者<Approved by>
(PASS)	()	()

No.3291-2