

SAMPLE

検査成績表

<TEST REPORT>

1/1

品名<Model Name> (温度ハイテスタ<TEMPERATURE HiTESTER>)
形名<Model Number> (3414)
製造番号<Serial No.> (No. 110712345)
検査年月日<Test Date> (2011-02-03)
(<YYYY-MM-DD>)
検査条件<Test Conditions> (23.0 °C, 50 %rh)

項目 <Item>	標準器指示値 <Standard>	許容範囲 <Tolerance>	計器指示値 <Indicated Value>
確度<Accuracy>			
	-50.0 °C	-51.1 °C ~ -48.9 °C	(-48.9 °C)
	0.0 °C	-1.0 °C ~ 1.0 °C	(1.0 °C)
	20.0 °C	19.0 °C ~ 21.0 °C	(21.0 °C)
	40.0 °C	38.9 °C ~ 41.1 °C	(41.1 °C)
	60.0 °C	58.9 °C ~ 61.1 °C	(61.1 °C)
	80.0 °C	78.8 °C ~ 81.2 °C	(81.2 °C)
	100.0 °C	98.7 °C ~ 101.3 °C	(101.3 °C)
	120.0 °C	118.7 °C ~ 121.3 °C	(121.0 °C)
	140.0 °C	138.6 °C ~ 141.4 °C	(141.4 °C)
	160.0 °C	158.6 °C ~ 161.4 °C	(161.4 °C)
	180.0 °C	178.5 °C ~ 181.5 °C	(181.0 °C)
	190.0 °C	188.5 °C ~ 191.5 °C	(191.5 °C)

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected by>	承認者<Approved by>
(PASS)	()	()