

品名<Model Name> (デジタルメグオームハイテスタ <DIGITAL MΩ HiTESTER>)
 形名<Model Number> (3453)
 製造番号<Serial No.> (No. 161202123)
 検査年月日<Test Date> (2016-12-02)
 <YYYY-MM-DD>
 検査条件<Test Conditions> (25.0 °C, 55 %rh)

項目 <Item>	ファンクション <Function>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>				表示値 <Indicated Value>
確度 <Accuracy>	Ω		0.9 Ω	0.1	Ω	~	1.7	Ω (0.9 Ω)
			360 Ω	352.0	Ω	~	368.0	Ω (360.0 Ω)
	~V		0V	0	V	~	8	V (0 V)
			540V(60Hz)	516	V	~	564	V (540 V)
	MΩ 125V	4MΩ	0 MΩ	0.000	MΩ	~	0.006	MΩ (0.000 MΩ)
		4MΩ	0.1MΩ	0.095	MΩ	~	0.105	MΩ (0.100 MΩ)
		4MΩ	0.5MΩ	0.487	MΩ	~	0.513	MΩ (0.500 MΩ)
		40MΩ	10 MΩ	9.77	MΩ	~	10.23	MΩ (10.00 MΩ)
		40MΩ	20 MΩ	19.00	MΩ	~	21.00	MΩ (20.00 MΩ)
	MΩ 250V	4MΩ	0 MΩ	0.000	MΩ	~	0.006	MΩ (0.000 MΩ)
		4MΩ	0.2MΩ	0.193	MΩ	~	0.207	MΩ (0.200 MΩ)
		40MΩ	20 MΩ	19.57	MΩ	~	20.43	MΩ (20.00 MΩ)
		400MΩ	50 MΩ	47.5	MΩ	~	52.5	MΩ (50.0 MΩ)
		2000MΩ	1900 MΩ	1805	MΩ	~	1995	MΩ (1900 MΩ)
	MΩ 500V	4MΩ	0 MΩ	0.000	MΩ	~	0.006	MΩ (0.000 MΩ)
		4MΩ	0.2MΩ	0.193	MΩ	~	0.207	MΩ (0.200 MΩ)
		400MΩ	50 MΩ	48.7	MΩ	~	51.3	MΩ (50.0 MΩ)
		2000MΩ	1900 MΩ	1805	MΩ	~	1995	MΩ (1900 MΩ)
	MΩ 1000V	4MΩ	0 MΩ	0.000	MΩ	~	0.006	MΩ (0.000 MΩ)
		4MΩ	0.2MΩ	0.193	MΩ	~	0.207	MΩ (0.200 MΩ)
		400MΩ	100 MΩ	97.7	MΩ	~	102.3	MΩ (100.0 MΩ)
		4000MΩ	999 MΩ	977	MΩ	~	1021	MΩ (999 MΩ)
		4000MΩ	3800 MΩ	3610	MΩ	~	3990	MΩ (3800 MΩ)
		4MΩ	3.3MΩ	3.231	MΩ	~	3.369	MΩ (3.300 MΩ)
		400MΩ	330 MΩ	323.1	MΩ	~	336.9	MΩ (330.0 MΩ)

項目 <Item>	ファンクション <Function>	負荷抵抗 <Load Resistance>	許容範囲 <Tolerance>				出力電圧 <Output Voltage>
出力電圧確度 <Output Voltage Accuracy>							
	MΩ 125V	100.5 MΩ	125	V	~	150	V (125 V)
	MΩ 250V	100.5 MΩ	250	V	~	300	V (250 V)
	MΩ 500V	100.5 MΩ	500	V	~	600	V (500 V)
	MΩ 1000V	100.5 MΩ	1000	V	~	1200	V (1000 V)
	MΩ 500V	0.5 MΩ	500	V	~	600	V (500 V)
	MΩ 1000V	2 MΩ	1000	V	~	1200	V (1000 V)

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected By>	承認者<Approved By>
(PASS)	()	()