

SAMPLE

検査成績表

<TEST REPORT>

HIOKI
HIOKI E.E. CORPORATION

1/1

品名<Model Name> (デジタルメグオームハイテスタ <DIGITAL MΩ HiTESTER>)
 形名<Model Number> (3453-01)
 製造番号<Serial No.> (No. 061234567)
 検査年月日<Test Date> (2018-04-11)
 <YYYY-MM-DD>
 検査条件<Test Conditions> (23.0 °C, 50 %rh)

項目 <Item>	ファンクション <Function>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>				表示値 <Indicated Value>
確度 <Accuracy>	Ω		0.9 Ω	0.1 Ω	~	1.7 Ω	(0.9 Ω)	
			360 Ω	352.0 Ω	~	368.0 Ω	(360.0 Ω)	
	~V		0V	0 V	~	8 V	(0 V)	
			540V(60Hz)	516 V	~	564 V	(540 V)	
MΩ 125V		4MΩ	0 MΩ	0.000 MΩ	~	0.006 MΩ	(0.000 MΩ)	
		4MΩ	0.1MΩ	0.095 MΩ	~	0.105 MΩ	(0.100 MΩ)	
		4MΩ	0.5MΩ	0.487 MΩ	~	0.513 MΩ	(0.500 MΩ)	
		40MΩ	10 MΩ	9.77 MΩ	~	10.23 MΩ	(10.00 MΩ)	
		40MΩ	20 MΩ	19.00 MΩ	~	21.00 MΩ	(20.00 MΩ)	
MΩ 250V		4MΩ	0 MΩ	0.000 MΩ	~	0.006 MΩ	(0.000 MΩ)	
		4MΩ	0.2MΩ	0.193 MΩ	~	0.207 MΩ	(0.200 MΩ)	
		40MΩ	20 MΩ	19.57 MΩ	~	20.43 MΩ	(20.00 MΩ)	
		400MΩ	50 MΩ	47.5 MΩ	~	52.5 MΩ	(50.0 MΩ)	
		2000MΩ	1900 MΩ	1805 MΩ	~	1995 MΩ	(1900 MΩ)	
MΩ 500V		4MΩ	0 MΩ	0.000 MΩ	~	0.006 MΩ	(0.000 MΩ)	
		4MΩ	0.2MΩ	0.193 MΩ	~	0.207 MΩ	(0.200 MΩ)	
		400MΩ	50 MΩ	48.7 MΩ	~	51.3 MΩ	(50.0 MΩ)	
		2000MΩ	1900 MΩ	1805 MΩ	~	1995 MΩ	(1900 MΩ)	
MΩ 1000V		4MΩ	0 MΩ	0.000 MΩ	~	0.006 MΩ	(0.000 MΩ)	
		4MΩ	0.2MΩ	0.193 MΩ	~	0.207 MΩ	(0.200 MΩ)	
		400MΩ	100 MΩ	97.7 MΩ	~	102.3 MΩ	(100.0 MΩ)	
		4000MΩ	999 MΩ	977 MΩ	~	1021 MΩ	(999 MΩ)	
		4000MΩ	3800 MΩ	3610 MΩ	~	3990 MΩ	(3800 MΩ)	
		4MΩ	3.3MΩ	3.231 MΩ	~	3.369 MΩ	(3.300 MΩ)	
		400MΩ	330 MΩ	323.1 MΩ	~	336.9 MΩ	(330.0 MΩ)	

項目 <Item>	ファンクション <Function>	負荷抵抗 <Load Resistance>	許容範囲 <Tolerance>				出力電圧 <Output Voltage>
出力電圧確度 <Output Voltage Accuracy>							
MΩ 125V		100.5 MΩ	125 V	~	150 V	(125 V)	
MΩ 250V		100.5 MΩ	250 V	~	300 V	(250 V)	
MΩ 500V		100.5 MΩ	500 V	~	600 V	(500 V)	
MΩ 1000V		100.5 MΩ	1000 V	~	1200 V	(1000 V)	
MΩ 500V		0.5 MΩ	500 V	~	600 V	(500 V)	
MΩ 1000V		2 MΩ	1000 V	~	1200 V	(1000 V)	

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected By>	承認者<Approved By>
(PASS)	()	()

No.3453-10