

SAMPLE

検査成績表

<TEST REPORT>

1/1

品名<Model Name> (デジタルメグオームハイテスタ <DIGITAL MΩ HiTESTER>)
 形名<Model Number> (3454-11)
 製造番号<Serial No.> (No. 161202123)
 検査年月日<Test Date> (2016-12-02)
 (<YYYY-MM-DD>)
 検査条件<Test Conditions> (25.0 °C, 55 %rh)

項目 <Item>	ファンクション <Function>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>	表示値 <Indicated Value>
確度 <Accuracy>	Ω	40 Ω	0.2 Ω	0.14 Ω ~ 0.26 Ω	(0.20 Ω)
		40 Ω	36 Ω	34.86 Ω ~ 37.14 Ω	(36.00 Ω)
		400 Ω	360 Ω	348.6 Ω ~ 371.4 Ω	(360.0 Ω)
		4k Ω	3.6k Ω	3.486 kΩ ~ 3.714 kΩ	(3.600 kΩ)
		40k Ω	36k Ω	34.86 kΩ ~ 37.14 kΩ	(36.00 kΩ)
		400k Ω	360k Ω	341.4 kΩ ~ 378.6 kΩ	(360.0 kΩ)
		4M Ω	3.6M Ω	3.414 MΩ ~ 3.786 MΩ	(3.600 MΩ)
	ACV		0V	-6 V ~ 6 V	(0 V)
			540V(50Hz)	518 V ~ 562 V	(540 V)
250V		4M Ω	0 M Ω	0.000 MΩ ~ 0.005 MΩ	(0.000 MΩ)
		4M Ω	0.2M Ω	0.190 MΩ ~ 0.210 MΩ	(0.200 MΩ)
		40M Ω	10 M Ω	9.66 MΩ ~ 10.34 MΩ	(10.00 MΩ)
		400M Ω	50 M Ω	48.1 MΩ ~ 51.9 MΩ	(50.0 MΩ)
		500M Ω	450 M Ω	423 MΩ ~ 477 MΩ	(450 MΩ)
500V		4M Ω	0 M Ω	0.000 MΩ ~ 0.005 MΩ	(0.000 MΩ)
		4M Ω	0.2M Ω	0.190 MΩ ~ 0.210 MΩ	(0.200 MΩ)
		40M Ω	10 M Ω	9.66 MΩ ~ 10.34 MΩ	(10.00 MΩ)
		400M Ω	50 M Ω	48.1 MΩ ~ 51.9 MΩ	(50.0 MΩ)
		500M Ω	450 M Ω	423 MΩ ~ 477 MΩ	(450 MΩ)
1000V		4M Ω	0 M Ω	0.000 MΩ ~ 0.005 MΩ	(0.000 MΩ)
		4M Ω	0.2M Ω	0.190 MΩ ~ 0.210 MΩ	(0.200 MΩ)
		400M Ω	100 M Ω	96.6 MΩ ~ 103.4 MΩ	(100.0 MΩ)
		400M Ω	200 M Ω	193.6 MΩ ~ 206.4 MΩ	(200.0 MΩ)
		4000M Ω	3600 M Ω	3415 MΩ ~ 3785 MΩ	(3600 MΩ)
		4M Ω	3.6M Ω	3.488 MΩ ~ 3.712 MΩ	(3.600 MΩ)
		40M Ω	36 M Ω	34.88 MΩ ~ 37.12 MΩ	(36.00 MΩ)
		400M Ω	360 M Ω	341.5 MΩ ~ 378.5 MΩ	(360.0 MΩ)

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected By>	承認者<Approved By>
(PASS)	()	()

No.345411-9