

SAMPLE

検査成績表

<TEST REPORT>

1/1

品名<Model Name> (ACミリオームハイテスタ <ACmΩ HiTESTER>)
 形名<Model Number> (3560)
 製造番号<Serial No.> (No. 030912611)
 検査年月日<Test Date> (2003-09-24)
 (<YYYY-MM-DD>)
 検査条件<Test Conditions> (25.1 °C, 48 %rh)
 電源周波数<Supply Frequency> (60 Hz)

項目 <Item>	レンジ <Range>	入力 <Input>	(標準器校正値) <Std. Cal. Value>	許容範囲 <Tolerance>	表示値 <Indicated Value>
確度<Accuracy>					
Sample Rate SLOW	3kΩ	3.0000kΩ (	3.0000 kΩ)	2.9842 kΩ ~	3.0158 kΩ (
		0.3000kΩ (	0.3000 kΩ)	0.2977 kΩ ~	0.3023 kΩ (
	300 Ω	300.00 Ω (	299.98 Ω)	298.40 Ω ~	301.56 Ω (
	30 Ω	30.000 Ω (	30.006 Ω)	29.848 Ω ~	30.164 Ω (
	3 Ω	3.0000 Ω (	3.0008 Ω)	2.9850 Ω ~	3.0166 Ω (
	300mΩ	300.00mΩ (	300.27 mΩ)	298.69 mΩ ~	301.85 mΩ (
	30mΩ	30.000mΩ (	30.067 mΩ)	29.909 mΩ ~	30.225 mΩ (
	50 V	49.000 V		48.971 V ~	49.029 V (
	5 V	4.9000 V		4.8971 V ~	4.9029 V (
Sample Rate MEDIUM	3kΩ	3.0000kΩ (	3.0000 kΩ)	2.9839 kΩ ~	3.0161 kΩ (
	30mΩ	30.000mΩ (	30.067 mΩ)	29.906 mΩ ~	30.228 mΩ (
Sample Rate FAST	3kΩ	3.000kΩ (	3.000 kΩ)	2.979 kΩ ~	3.021 kΩ (
	30mΩ	30.00mΩ (	30.07 mΩ)	29.84 mΩ ~	30.30 mΩ (
	5 V	4.9000 V		4.8966 V ~	4.9034 V (

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

標準器校正値を使用しているポイントの許容範囲は、標準器校正値を基準に定めています。

<The tolerance for each point using the standard calibration value is based on the standard calibration value.>

総合判定<Overall Result>	検査者<Inspected by>	承認者<Approved by>
(PASS)	()	()

No.3560-3