

SAMPLE

検査成績表

<TEST REPORT>

1/1

品名<Model Name> (パルスロガー<PULSE LOGGER>)
 形名<Model Number> (3672)
 製造番号<Serial No.> (No.110812345)
 検査年月日<Test Date> (2011-08-02)
 <YYYY-MM-DD>
 検査条件<Test Conditions> (23.0 °C, 50 %rh)

項目 <Item>	入力 <Input>	許容範囲 <Tolerance>	表示値 <Indicated Value>
測定確度<Measurement accuracy>			
-1.機械式接点入力 (Filter:ON) * <Mechanical point of contact input>	10Hz(0-2V矩形波) <Wave:0-2V Square>	49 ~ 51	(50)
-2.電子式接点入力 (Filter:OFF) <Electronic type point of contact input>	1kHz(0-2V矩形波) <Wave:0-2V Square>	999 ~ 1001	(1000)

備考<Note>

*測定条件:記録間隔5秒

<Measurement conditions : Record interval 5 seconds>

FAIL判定箇所は、グレー表示としています。

<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected by>	承認者<Approved by>
(PASS)	()	()