

SAMPLE

検査成績表

<TEST REPORT>

1/1

品名<Model Name> (クランプロガー<CLAMP LOGGER>)
形名<Model Number> (3673)
製造番号<Serial No.> (No.110812345)
検査年月日<Test Date> (2011-08-03)
<YYYY-MM-DD>
検査条件<Test Conditions> (23.0 °C, 50 %rh)

項目 <Item>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>	表示値 <Indicated Value>
測定確度 <Measurement accuracy>				
AC電流 <AC Current>	50 A	0 A (DC)	0.00 A ~ 0.05 A	CH1 (0.00 A) CH2 (0.00 A)
		3 A (60Hz)	2.92 A ~ 3.08 A	(3.00 A) (3.00 A)
		10 A (60Hz)	9.85 A ~ 10.15 A	(10.00 A) (10.00 A)
		30 A (60Hz)	29.65 A ~ 30.35 A	(30.00 A) (30.00 A)
	500 A	0 A (DC)	0.0 A ~ 0.5 A	(0.0 A) (0.0 A)
		30 A (60Hz)	29.2 A ~ 30.8 A	(30.0 A) (30.0 A)
		100 A (60Hz)	98.5 A ~ 101.5 A	(100.0 A) (100.0 A)
		300 A (60Hz)	296.5 A ~ 303.5 A	(300.0 A) (300.0 A)

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected by>	承認者<Approved by>
(PASS)	()	()