

SAMPLE

検査成績表

<TEST REPORT>

1/1

品名<Model Name> (AC電圧ロガー<AC VOLTAGE LOGGER>)
 形名<Model Number> (3674)
 製造番号<Serial No.> (No.110812345)
 検査年月日<Test Date> (2011-08-02)
 <YYYY-MM-DD>
 検査条件<Test Conditions> (23.0 °C, 50 %rh)

項目 <Item>	入力 <Input>	許容範囲 <Tolerance>	表示値 <Indicated Value>
測定確度<Measurement Accuracy>			
AC電圧<AC Voltage>	0V (DC)	0.0 V ~ 0.5 V	(0.0 V)
	30V (60Hz)	29.2 V ~ 30.8 V	(30.0 V)
	100V (60Hz)	98.5 V ~ 101.5 V	(100.0 V)
	400V (60Hz)	395.5 V ~ 404.5 V	(400.0 V)

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected by>	承認者<Approved by>
(PASS)	()	()