

SAMPLE

検査成績表

<TEST REPORT>

HIOKI
HIOKI ELECTRIC CORPORATION

1/2

品名<Model Name> (マイクロハイコーダ<MICRO HiCORDER>)
形名<Model Number> (8205)
製造番号<Serial No.> (No.110812345)
検査年月日<Test Date> (2011-08-04)
<YYYY-MM-DD>
検査条件<Test Conditions> (23.0 °C, 50 %rh)
電源周波数<Supply Frequency> (60Hz)

項目 <Item>	レンジ <Range>	試験点 <Test Point>	許容範囲 <Tolerance>	標準器指示値 <Standard>
--------------	----------------	---------------------	---------------------	----------------------

確度<Accuracy>

DC V

0.1 V	0.1 V	0.0980 V	~ 0.1020 V	(0.1000 V)
0.2 V	0.2 V	0.1960 V	~ 0.2040 V	(0.2000 V)
0.5 V	0.5 V	0.490 V	~ 0.510 V	(0.500 V)
1 V	1 V	0.980 V	~ 1.020 V	(1.000 V)
2 V	2 V	1.960 V	~ 2.040 V	(2.000 V)
5 V	5 V	4.90 V	~ 5.10 V	(5.00 V)
10 V	10 V	9.80 V	~ 10.20 V	(10.00 V)
20 V	20 V	19.60 V	~ 20.40 V	(20.00 V)
50 V	50 V	49.0 V	~ 51.0 V	(50.0 V)
100 V	100 V	98.0 V	~ 102.0 V	(100.0 V)
200 V	200 V	196.0 V	~ 204.0 V	(200.0 V)
500 V	500 V	490 V	~ 510 V	(500 V)
10 V	-8 V	-8.20 V	~ -7.80 V	(-8.00 V)

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected by>	承認者<Approved by>
(PASS)	()	()

SAMPLE 検査成績表

<TEST REPORT>

HIOKI
HIOKI E.E. CORPORATION

2/2

製造番号<Serial No.> (No.110812345)
検査年月日<Test Date> (2011-08-04)
<YYYY-MM-DD>

項目 <Item>	レンジ <Range>	試験点 <Test Point>	許容範囲 <Tolerance>	標準器指示値 <Standard>
--------------	----------------	---------------------	---------------------	----------------------

確度<Accuracy>

AC V

0.1 V	0.1 V/50Hz	0.0980 V	~ 0.1020 V	(0.1020 V)
0.2 V	0.2 V/50Hz	0.1960 V	~ 0.2040 V	(0.2040 V)
0.5 V	0.5 V/50Hz	0.490 V	~ 0.510 V	(0.510 V)
1 V	1 V/50Hz	0.980 V	~ 1.020 V	(1.020 V)
2 V	2 V/50Hz	1.960 V	~ 2.040 V	(2.040 V)
5 V	5 V/50Hz	4.90 V	~ 5.10 V	(5.10 V)
10 V	10 V/50Hz	9.80 V	~ 10.20 V	(10.20 V)
20 V	20 V/50Hz	19.60 V	~ 20.40 V	(20.40 V)
50 V	50 V/50Hz	49.0 V	~ 51.0 V	(51.0 V)
100 V	100 V/50Hz	98.0 V	~ 102.0 V	(102.0 V)
200 V	200 V/50Hz	196.0 V	~ 204.0 V	(204.0 V)
500 V	500 V/50Hz	490 V	~ 510 V	(510 V)

AC A

10 A	10 mA/50Hz	9.80 mA	~ 10.20 mA	(10.20 mA)
20 A	20 mA/50Hz	19.60 mA	~ 20.40 mA	(20.40 mA)
50 A	50 mA/50Hz	49.0 mA	~ 51.0 mA	(51.0 mA)
100 A	100 mA/50Hz	98.0 mA	~ 102.0 mA	(102.0 mA)
200 A	200 mA/50Hz	196.0 mA	~ 204.0 mA	(204.0 mA)
500 A	500 mA/50Hz	490 mA	~ 510 mA	(510 mA)

備考<Note>

AC Aの値は、各レンジの定格値の1/1000を示す。

<The value of AC A indicates 1/1000 of rated value for each ranges.>