

SAMPLE

検査成績表

<TEST REPORT>

HIOKI
HIOKI ELECTRONIC CORPORATION

1/1

品名<Model Name> (マイクロハイコーダ<MICRO HiCORDER>)
形名<Model Number> (8206)
製造番号<Serial No.> (No. 110812345)
検査年月日<Test Date> (2011-08-18)
(YYYY-MM-DD)
検査条件<Test Conditions> (23.0 °C, 50 %rh)
電源周波数<Supply Frequency> (60 Hz)

項目 <Item>	レンジ <Range>	試験点 <Test Point>	許容範囲 <Tolerance>	標準器指示値 <Standard>
--------------	----------------	---------------------	---------------------	----------------------

確度<Accuracy>

AC V

100 V	100 V/50Hz	97.5 V ~ 102.5 V	(100.0 V)
200 V	200 V/50Hz	195.0 V ~ 205.0 V	(200.0 V)
500 V	500 V/50Hz	487.5 V ~ 512.5 V	(500.0 V)

AC A

10 A	10 mA/50Hz	9.80 mA ~ 10.20 mA	(10.00 mA)
20 A	20 mA/50Hz	19.60 mA ~ 20.40 mA	(20.00 mA)
50 A	50 mA/50Hz	49.0 mA ~ 51.0 mA	(50.0 mA)
100 A	100 mA/50Hz	98.0 mA ~ 102.0 mA	(100.0 mA)
200 A	200 mA/50Hz	196.0 mA ~ 204.0 mA	(200.0 mA)
500 A	500 mA/50Hz	490 mA ~ 510 mA	(500 mA)

備考<Note>

AC Aの値は、各レンジの定格値の1/1000を示す。

<The value of AC A indicates 1/1000 of rated value for each ranges.>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected by>	承認者<Approved by>
(PASS)	()	()