

SAMPLE

検査成績表
<TEST REPORT>

1/1

品名<Model Name> (DC/RMSユニット<DC/RMS UNIT>)
形名<Model Number> (8917)
製造番号<Serial No.> (No.111234567)
検査年月日<Test Date> (2011-11-03)
<YYYY-MM-DD>
検査条件<Test Conditions> (23.0 °C, 60 %rh)

項目 <Item>	レンジ <Range>	試験点 <Test Point>	許容範囲 <Tolerance>	測定値 <Measured Value>
1.ゼロポジション確度 <Zero Position Accuracy>				
	5 mV/DIV	0 mV	± 0.100 mV	(0.063 mV)
2.DC MODE 振幅確度 <DC MODE Accuracy>				
	5 mV/DIV	DC 100 mV	± 0.30 mV	(100.06 mV)
	10 mV/DIV	DC 200 mV	± 0.60 mV	(200.13 mV)
	20 mV/DIV	DC 400 mV	± 1.20 mV	(400.25 mV)
	50 mV/DIV	DC 1 V	± 0.0030 V	(1.0006 V)
	100 mV/DIV	DC 2 V	± 0.0060 V	(2.0013 V)
	200 mV/DIV	DC 4 V	± 0.0120 V	(4.0025 V)
	500 mV/DIV	DC 10 V	± 0.030 V	(10.006 V)
	1 V/DIV	DC 20 V	± 0.060 V	(20.013 V)
	2 V/DIV	DC 40 V	± 0.120 V	(40.025 V)
	5 V/DIV	DC 100 V	± 0.30 V	(100.06 V)
	10 V/DIV	DC 200 V	± 0.60 V	(200.13 V)
	20 V/DIV	DC 350 V	± 1.20 V	(350.22 V)
3.RMS MODE 振幅確度 <RMS MODE Accuracy>				
	5 mV/DIV	AC50Hz 100 mV	± 1.00 mV	(100.06 mV)
	10 mV/DIV	AC50Hz 200 mV	± 2.00 mV	(200.13 mV)
	20 mV/DIV	AC50Hz 400 mV	± 4.00 mV	(400.25 mV)
	50 mV/DIV	AC50Hz 1 V	± 0.0100 V	(1.0006 V)
	100 mV/DIV	AC50Hz 2 V	± 0.0200 V	(2.0013 V)
	200 mV/DIV	AC50Hz 4 V	± 0.0400 V	(4.0025 V)
	500 mV/DIV	AC50Hz 10 V	± 0.100 V	(10.006 V)
	1 V/DIV	AC50Hz 20 V	± 0.200 V	(20.013 V)
	2 V/DIV	AC50Hz 40 V	± 0.400 V	(40.025 V)
	5 V/DIV	AC50Hz 100 V	± 1.00 V	(100.06 V)
	10 V/DIV	AC50Hz 200 V	± 2.00 V	(200.13 V)
	20 V/DIV	AC50Hz 350 V	± 4.00 V	(350.22 V)

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>

(PASS)

検査者<Inspected by>

()

承認者<Approved by>

()

No.8917-5