

SAMPLE

検査成績表

<TEST REPORT>

1/1

品名<Model Name> (温度ユニット<TEMPERATURE UNIT>)
 形名<Model Number> (8935)
 製造番号<Serial No.> (No. 110812345)
 検査年月日<Test Date> (2011-08-04)
 <YYYY-MM-DD>
 検査条件<Test Conditions> (23.0 °C, 50 %rh)
 電源周波数<Supply Frequency> (60Hz)

項目 <Item>	レンジ <Range>	試験点 <Test Point>	許容範囲 <Tolerance>	測定結果 <Measured Value>
--------------	----------------	---------------------	---------------------	--------------------------

確度<Accuracy>

K

10 °C/DIV	90 °C	87.0°C ~ 93.0°C	(90.0 °C)
20 °C/DIV	180 °C	176.0°C ~ 184.0°C	(180.0 °C)
50 °C/DIV	450 °C	443.0°C ~ 457.0°C	(450.0 °C)
100 °C/DIV	900 °C	888.0°C ~ 912.0°C	(900.0 °C)

J

10 °C/DIV	90 °C	87.0°C ~ 93.0°C	(90.0 °C)
20 °C/DIV	180 °C	176.0°C ~ 184.0°C	(180.0 °C)
50 °C/DIV	450 °C	443.0°C ~ 457.0°C	(450.0 °C)

T

10 °C/DIV	90 °C	87.0°C ~ 93.0°C	(90.0 °C)
20 °C/DIV	180 °C	176.0°C ~ 184.0°C	(180.0 °C)

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected by>	承認者<Approved by>
(PASS)	()	()

No.8935-5