HIOKI
HIOKI E.E. CORPORATION

SAMPLE

WT-8771 INSPECTION DATA SHEET

Model Number	WT-8771	Inspected By	
Serial No.	123456789		
Test Data	2014-06-12 <YYYY-MM-DD>	Approved By	
Test Conditions	24.0 ℃、 60 %rh		

1. AC Withstanding voltage meter

Voltage meter accuracy : ±(1% of reading +5digits)
Current meter accuracy : ＜1.2kV ± (1.5% of reading+5digits)
≧1.2kV ± (1.5% of reading+5digits/kV)

AC1000V

Set voltage(kV)	Indicated voltage(kV)	Measurement voltage(V)
0.100	0.100	100.0
1.000	1.000	1000
5.000	5.000	5000

Indicated current(mA)	Measurement current(mA)
0.500	0.500
5.00	5.00
20.00	20.00

Judge PASS