

SAMPLE

WT-8772 INSPECTION DATA SHEET

HIOKI
HIOKI E.E. CORPORATION

Model Number	WT-8772	Inspected By	
Serial No.	123456789	Approved By	
Test Data	2014-06-12		
	<YYYY-MM-DD>		
Test Conditions	24.0 °C、	60 %rh	

1. AC Withstanding voltage meter

Voltage meter accuracy : $\pm$ (1% of reading +5digits)
Current meter accuracy : $< 1.2\text{kV} \pm (1.5\% \text{ of reading} + 5\text{digits})$
 $\geq 1.2\text{kV} \pm (1.5\% \text{ of reading} + 5\text{digits}/\text{kV})$

AC1000V

Set voltage(kV)	Indicated voltage(kV)	Measurement voltage(V)
0.100	0.100	100.0
1.000	1.000	998
5.000	5.000	5000

Indicated current(mA)	Measurement current(mA)
0.500	0.500
5.00	5.00
20.00	20.00

Judge PASS

2. DC Withstanding voltage meter

Voltage meter accuracy : $\pm$ (1% of reading +5digits)
Current meter accuracy : $< 1.2\text{kV} \pm (1.5\% \text{ of reading} + 5\text{digits})$
 $\geq 1.2\text{kV} \pm (1.5\% \text{ of reading} + 2\text{digits}/\text{kV})$

DC1000V

Set voltage(kV)	Indicated voltage(kV)	Measurement voltage(V)
0.100	0.100	100.0
1.000	1.000	999
6.000	6.000	6000

Indicated current(mA)	Measurement current(mA)
0.500	0.500
0.996	0.995
4.91	4.92

Judge PASS

No.WT8772E-2