

SAMPLE

検査成績表
<TEST REPORT>HIOKI
HIOKI E.E. CORPORATION

1/1

品名<Model Name> (ストレインユニット<STRAIN UNIT>)
 形名<Model Number> (8939)
 製造番号<Serial No.> (No. 1998-1040948)
 検査年月日<Test Date> (2011-06-01)
 <YYYY-MM-DD>
 検査条件<Test Conditions> (23.6 °C, 52 %rh)

項目 <Item>	レンジ <Range>	検査点 <Inspection Point>	許容範囲 <Tolerance>	測定値 <Measured Value>
--------------	----------------	---------------------------	---------------------	-------------------------

1.DC振幅確度<DC Amplitude Accuracy>

f.s.=10DIV の製品に実装時 <f.s. on recorder = 10DIV>

CH1

CH2

50 $\mu\epsilon$ /DIV	500 $\mu\epsilon$	495.50 $\mu\epsilon$ ~ 504.50 $\mu\epsilon$ (499.06 $\mu\epsilon$) (499.69 $\mu\epsilon$)
100 $\mu\epsilon$ /DIV	1000 $\mu\epsilon$	993.00 $\mu\epsilon$ ~ 1007.00 $\mu\epsilon$ (998.75 $\mu\epsilon$) (998.75 $\mu\epsilon$)
200 $\mu\epsilon$ /DIV	2000 $\mu\epsilon$	1988.0 $\mu\epsilon$ ~ 2012.0 $\mu\epsilon$ (1995.0 $\mu\epsilon$) (1998.8 $\mu\epsilon$)
500 $\mu\epsilon$ /DIV	5000 $\mu\epsilon$	4973.0 $\mu\epsilon$ ~ 5027.0 $\mu\epsilon$ (4993.8 $\mu\epsilon$) (4993.8 $\mu\epsilon$)
1000 $\mu\epsilon$ /DIV	10000 $\mu\epsilon$	9948.0 $\mu\epsilon$ ~ 10052.0 $\mu\epsilon$ (9981.3 $\mu\epsilon$) (9981.3 $\mu\epsilon$)
2000 $\mu\epsilon$ /DIV	20000 $\mu\epsilon$	19898.0 $\mu\epsilon$ ~ 20102.0 $\mu\epsilon$ (19975.0 $\mu\epsilon$) (19988.0 $\mu\epsilon$)

f.s.=20DIV の製品に実装時 <f.s. on recorder = 20DIV>

20 $\mu\epsilon$ /DIV	400 $\mu\epsilon$	396.00 $\mu\epsilon$ ~ 404.00 $\mu\epsilon$ (399.25 $\mu\epsilon$) (399.00 $\mu\epsilon$)
50 $\mu\epsilon$ /DIV	1000 $\mu\epsilon$	993.00 $\mu\epsilon$ ~ 1007.00 $\mu\epsilon$ (998.75 $\mu\epsilon$) (998.75 $\mu\epsilon$)
100 $\mu\epsilon$ /DIV	2000 $\mu\epsilon$	1988.0 $\mu\epsilon$ ~ 2012.0 $\mu\epsilon$ (1995.0 $\mu\epsilon$) (1998.8 $\mu\epsilon$)
200 $\mu\epsilon$ /DIV	4000 $\mu\epsilon$	3978.0 $\mu\epsilon$ ~ 4022.0 $\mu\epsilon$ (3992.5 $\mu\epsilon$) (3997.5 $\mu\epsilon$)
500 $\mu\epsilon$ /DIV	10000 $\mu\epsilon$	9948.0 $\mu\epsilon$ ~ 10052.0 $\mu\epsilon$ (9981.3 $\mu\epsilon$) (9981.3 $\mu\epsilon$)
1000 $\mu\epsilon$ /DIV	20000 $\mu\epsilon$	19898.0 $\mu\epsilon$ ~ 20102.0 $\mu\epsilon$ (19975.0 $\mu\epsilon$) (19988.0 $\mu\epsilon$)

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result> (PASS)	検査者<Inspected by> ()	承認者<Approved by> ()
----------------------------------	--------------------------	-------------------------