

SAMPLE

検査成績表

<TEST REPORT>

1/1

品名<Model Name> (アナログユニット<ANALOG UNIT>)
 形名<Model Number> (8944)
 製造番号<Serial No.> (No.)
 検査年月日<Test Date> ()
 <YYYY-MM-DD>
 検査条件<Test Conditions> (°C, %rh)

項目 <Item>	レンジ <Range>	試験点 <Test Point>	許容範囲 <Tolerance>	測定結果 <Measured Value>
ゼロポジション確度<Zero Position>				
	10 mV/DIV	0 mV	± 0.100 mV	(0.099 mV)
確度<Accuracy>				
	10 mV/DIV	100 mV	± 0.25 mV	(100.25 mV)
	20 mV/DIV	200 mV	± 0.50 mV	(200.49 mV)
	50 mV/DIV	500 mV	± 1.25 mV	(501.24 mV)
	100 mV/DIV	1 V	± 0.0025 V	(1.0025 V)
	200 mV/DIV	2 V	± 0.0050 V	(2.0049 V)
	500 mV/DIV	5 V	± 0.0125 V	(5.0124 V)
	1 V/DIV	10 V	± 0.025 V	(10.025 V)
	2 V/DIV	20 V	± 0.050 V	(20.049 V)
	5 V/DIV	50 V	± 0.13 V	(50.124 V)
	10 V/DIV	100 V	± 0.25 V	(100.25 V)
	20 V/DIV	200 V	± 0.50 V	(200.49 V)
	50 V/DIV	350 V	± 1.250 V	(351.24 V)

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected by>	承認者<Approved by>
(PASS)	()	()

No8944-3