

SAMPLE

検査成績表

<TEST REPORT>

HIOKI
HIOKI E. E. CORPORATION

1/1

品名<Model Name> (DC/RMSユニット<DC/RMS UNIT>)
形名<Model Number> (8952)
製造番号<Serial No.> (No.)
検査年月日<Test Date> ()
検査条件<Test Conditions> (<YYYY-MM-DD>
22.4 °C, 50 %rh)

項目 <Item>	レンジ <Range>	検査点 <Test Point>	許容範囲 <Tolerance>	測定値 <Measured Value>
--------------	----------------	---------------------	---------------------	-------------------------

ゼロポジション確度<Zero Position Accuracy>(ADJ ON)

100 mV/DIV 0 mV -2.00 mV ~ 2.00 mV (0.50 mV)
*1

DC振幅確度<DC Amplitude Accuracy>(ADJ ON)

5 mV/DIV	100 mV	99.600 mV ~	100.400 mV	(100.069 mV)
10 mV/DIV	200 mV	199.200 mV ~	200.800 mV	(200.031 mV)
20 mV/DIV	400 mV	398.400 mV ~	401.600 mV	(400.338 mV)
50 mV/DIV	1 V	0.996 V ~	1.004 V	(1.001 V)
100 mV/DIV	2 V	1.992 V ~	2.008 V	(2.001 V)
200 mV/DIV	4 V	3.984 V ~	4.016 V	(4.001 V)
500 mV/DIV	10 V	9.960 V ~	10.040 V	(10.001 V)
1 V/DIV	20 V	19.920 V ~	20.080 V	(19.998 V)
2 V/DIV	40 V	39.840 V ~	40.160 V	(40.016 V)
5 V/DIV	100 V	99.600 V ~	100.400 V	(100.088 V)
10 V/DIV	200 V	199.200 V ~	200.800 V	(200.081 V)
20 V/DIV	400 V	398.400 V ~	401.600 V	(400.150 V)

RMS振幅確度<RMS Amplitude Accuracy>(ADJ ON)

5 mV/DIV	100 mV	98.000 mV ~	102.000 mV	(100.134 mV)
10 mV/DIV	200 mV	196.000 mV ~	204.000 mV	(200.275 mV)
20 mV/DIV	400 mV	392.000 mV ~	408.000 mV	(400.450 mV)
50 mV/DIV	1 V	0.980 V ~	1.020 V	(1.001 V)
100 mV/DIV	2 V	1.960 V ~	2.040 V	(2.001 V)
200 mV/DIV	4 V	3.920 V ~	4.080 V	(4.002 V)
500 mV/DIV	10 V	9.800 V ~	10.200 V	(10.012 V)
1 V/DIV	20 V	19.600 V ~	20.400 V	(20.022 V)
2 V/DIV	40 V	39.200 V ~	40.800 V	(40.061 V)
5 V/DIV	100 V	98.000 V ~	102.000 V	(100.172 V)
10 V/DIV	200 V	196.000 V ~	204.000 V	(199.963 V)
20 V/DIV	400 V	392.000 V ~	408.000 V	(400.088 V)

備考<Note>

- *1 ゼロポジション確度は、入力結合(カップリング)GNDにて校正しています。入力結合(カップリング)GNDとは内部回路で入力とGNDを短絡することです。<The accuracy of the zero position is calibrated by coupling GND. Coupling GND means to short input and GND at the inner circuit.>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected by>	承認者<Approved by>
(PASS)	()	()

No.8952-5