

SAMPLE

検査成績表

<TEST REPORT>

HIOKI
HIOKI E.E. CORPORATION

1/1

品名<Model Name> (高分解能ユニット<HIGH RESOLUTION UNIT>)
 形名<Model Number> (8953)
 製造番号<Serial No.> (No. 2001-0343024)
 検査年月日<Test Date> (2011-01-31)
 <YYYY-MM-DD>
 検査条件<Test Conditions> (22.4 °C, 45 %rh)

項目 <Item>	レンジ <Range>	検査点 <Test Point>	許容範囲 <Tolerance>	測定値 <Measured Value>
ゼロポジション確度<Zero Position Accuracy>(ADJ ON)				
	100 mV/DIV	0 mV	-2.00 mV ~ 2.00 mV	(-0.06 mV)
DC振幅確度<DC Amplitude Accuracy>(ADJ ON)				
5 mV/DIV	100 mV	99.800 mV ~ 100.200 mV	(99.984 mV)	
10 mV/DIV	200 mV	199.600 mV ~ 200.400 mV	(200.013 mV)	
20 mV/DIV	400 mV	399.200 mV ~ 400.800 mV	(400.038 mV)	
50 mV/DIV	1 V	0.998 V ~ 1.002 V	(1.000 V)	
100 mV/DIV	2 V	1.996 V ~ 2.004 V	(2.000 V)	
200 mV/DIV	4 V	3.992 V ~ 4.008 V	(4.001 V)	
500 mV/DIV	10 V	9.980 V ~ 10.020 V	(9.986 V)	
1 V/DIV	20 V	19.960 V ~ 20.040 V	(19.973 V)	
2 V/DIV	40 V	39.920 V ~ 40.080 V	(39.944 V)	
5 V/DIV	100 V	99.800 V ~ 100.200 V	(99.859 V)	
10 V/DIV	200 V	199.600 V ~ 200.400 V	(199.713 V)	
20 V/DIV	400 V	399.200 V ~ 400.800 V	(399.425 V)	

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected by>	承認者<Approved by>
(PASS)	()	()