

検査成績表
<TEST REPORT>

品名<Model Name> (ストレインユニット<STRAIN UNIT>)
 形名<Model Number> (8969)
 製造番号<Serial No.> (No. 091214106)
 検査年月日<Test Date> (2009-12-15)
 (<YYYY-MM-DD>)
 検査条件<Test Conditions> (22.7 °C, 52 %rh)

項目 <Item>	レンジ <Range>	検査点 <Test Point>	許容範囲 <Tolerance>	測定値 <Measured Value>
--------------	----------------	---------------------	---------------------	-------------------------

1. 測定精度<Measurement Accuracy>

				CH1	CH2
20 $\mu\epsilon$ /DIV	0 $\mu\epsilon$	-6.00 $\mu\epsilon$ ~	6.00 $\mu\epsilon$ (0.22 $\mu\epsilon$) (-0.08 $\mu\epsilon$)
	400 $\mu\epsilon$	394.00 $\mu\epsilon$ ~	406.00 $\mu\epsilon$ (399.82 $\mu\epsilon$) (399.84 $\mu\epsilon$)
50 $\mu\epsilon$ /DIV	0 $\mu\epsilon$	-9.00 $\mu\epsilon$ ~	9.00 $\mu\epsilon$ (0.12 $\mu\epsilon$) (-0.20 $\mu\epsilon$)
	1000 $\mu\epsilon$	991.00 $\mu\epsilon$ ~	1009.00 $\mu\epsilon$ (1000.08 $\mu\epsilon$) (1000.08 $\mu\epsilon$)
100 $\mu\epsilon$ /DIV	0 $\mu\epsilon$	-14.00 $\mu\epsilon$ ~	14.00 $\mu\epsilon$ (0.24 $\mu\epsilon$) (-0.24 $\mu\epsilon$)
	2000 $\mu\epsilon$	1986.00 $\mu\epsilon$ ~	2014.00 $\mu\epsilon$ (2000.40 $\mu\epsilon$) (1999.92 $\mu\epsilon$)
200 $\mu\epsilon$ /DIV	0 $\mu\epsilon$	-24.0 $\mu\epsilon$ ~	24.0 $\mu\epsilon$ (-0.2 $\mu\epsilon$) (-0.5 $\mu\epsilon$)
	4000 $\mu\epsilon$	3976.0 $\mu\epsilon$ ~	4024.0 $\mu\epsilon$ (4000.3 $\mu\epsilon$) (4000.2 $\mu\epsilon$)
500 $\mu\epsilon$ /DIV	0 $\mu\epsilon$	-54.0 $\mu\epsilon$ ~	54.0 $\mu\epsilon$ (-0.8 $\mu\epsilon$) (-0.8 $\mu\epsilon$)
	10000 $\mu\epsilon$	9946.0 $\mu\epsilon$ ~	10054.0 $\mu\epsilon$ (10000.8 $\mu\epsilon$) (10000.4 $\mu\epsilon$)
1000 $\mu\epsilon$ /DIV	0 $\mu\epsilon$	-104.0 $\mu\epsilon$ ~	104.0 $\mu\epsilon$ (-1.6 $\mu\epsilon$) (-1.6 $\mu\epsilon$)
	20000 $\mu\epsilon$	19896.0 $\mu\epsilon$ ~	20104.0 $\mu\epsilon$ (20002.4 $\mu\epsilon$) (20003.2 $\mu\epsilon$)

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result> (PASS)	検査者<Inspected by> ()	承認者<Approved by> ()
--------------------------------------	------------------------------	-----------------------------